

Invitation of quotation
for
Thermal Cyclers
At
Research Section,
All India Institute of Medical Sciences, Jodhpur

Inquiry No.: : AIIMS/RES(06)/2019/383

Inquiry Issue Date : September 13, 2019

Last Date of Submission : September 23, 2019 at 05:00 PM.



All India Institute of Medical Sciences, Jodhpur

Basni Phase - II, Jodhpur – 342005, Rajasthan

Telefax: 0291- 2740741, Extn. 3109, email: researchcell@aiimsjodhpur.edu.in

www.aiimsjodhpur.edu.in

**Invitation of quotation for Thermal Cyler at Research Section,
AIIMS Jodhpur**

Sealed Quotations are hereby invited by the undersigned on behalf of the Dean (Research), AIIMS Jodhpur for Supply of Thermal Cyler for the Institute as per terms & conditions mentioned below. The filled quotations along with all the required document must reach in the office of the undersigned on or before 23.09.2019 05.00 PM. The Envelope containing the quotation would please be sealed and super scribed as under:-

**“QUOTATION FOR SUPPLY OF THERMAL CYCLER AGAINST
INQUIRY NO. AIIMS/RES(06)/2019/383” DUE ON 23.09.2019 05.00 PM”**

1. Terms & Conditions:

- A) The quotations received after this deadline & unsealed shall not be entertained under any circumstances whatsoever. In case of postal delay this Institute will not be responsible. **The offer Submitted Fax/Email shall not be considered and no correspondence will be entertained in this matter.**
- B) Quotations must be in the enclosed prescribed Performa on the letter head of the firm duly signed by the Proprietor/ Partner/ Director or their authorized representative, In case of signing of quotation by the authorized representative letter of authorization must be attached with the quotation. Quotation must be submitted in **Research Section, AIIMS, Jodhpur.**
- C) Rates must be quoted in **Indian rupees** and as per the format specified taxes extra if any must be written separately.
- D) Rates must be quoted FOR basis (including Freight charges, Insurance, installation etc.)
- E) No overwriting or cutting is permitted in the rate. If found, the quotation shall be summarily rejected.
- F) The rates quoted must be valid for 60 days minimum from the date of opening of the quotation and silence of any tendered on this issue shall be treated as agreed with this condition.
- G) Becoming L1 will not be the criteria for awarding of purchase order unless the rates are reasonable & justified.
- H) RTGS/NEFT details need to be furnished by the supplier with the quotation on the letter head of supplier/firm/agency.
- I) The firm/agency may satisfy the following conditions and attach self-attested copy of the same with the quotation:
- Firm shall be registered with the Government of Rajasthan / Central Government.
 - The firm shall have valid GST/Other taxes and IT PAN.
 - **The firm should not be black listed by any Govt. Agency/Dept.**

INQUIRY NO. AIIMS/RES(06)/2019/383

- J) Quotations qualified by such vague and indefinite expressions such as “subject to prior confirmation”, “subject to immediate acceptance” etc. will be treated as vague offers and rejected accordingly. Any conditional quotation shall be rejected summarily.
- K) **Delivery Period** – within 30 days from Purchase order.
- L) **Liquidated Damage:** - If the supplier fails to deliver the material on or before the stipulated date, then a penalty at the rate of 0.5 % per week of the total order value shall be levied subject to maximum of 10% of the total order value.
- M) **Payment Terms:** Payment will be only after satisfactorily delivery / commissioning of material and after inspection by the AIIMS Jodhpur.
- N) **Disputes:** -In the event of any dispute or disagreement arising between the contractors and any other department of AIIMS Jodhpur with regards to the interpretation of “Terms & Conditions” of this inquiry, the same shall be referred to the Director, AIIMS Jodhpur whose decision will be final and binding upon the contractor.
- O) Research Section, AIIMS, Jodhpur reserves the right to increase or decrease quantity and / or amount of work. Decision of Quantity of material in the Research Section, AIIMS, Jodhpur will be final in this regard.
- P) Research Section, AIIMS, Jodhpur reserves the right to reject any quotation or part or the whole of inviting quotation process without assigning any reason. Decision of the Research Section, AIIMS, Jodhpur will be final in this regard.

2. Special Terms & Conditions:

- A) **Bidder must quote the product as per specification provided in Annexure 1.**
- B) **Catalog must be attached with quotation for technical evaluation.**
- C) **The Bidder must submit the GSTIN Registration and PAN Card self-attested copy with the quotation.**
- D) **The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.**

Sub Dean (Research)

Encl.: Annexure 1 (Specification)
Annexure 2 (Format of price bid)

Annexure 1

S. No	Particular	Specifications	Qty.
1.	Thermal Cycler	<ul style="list-style-type: none"> • System should be 96 well thermal cycler. • Should have ability to run at least six different assays at a time with independently controlled Peltier block. • System should be compatible for both 0.2ml tube and PCR plate. • Should provide standard and fast run mode (within 30 min.). • Ramping rate should not be less than 5.00⁰ C/sec. • Gradient technology should ensure identical ramp rates in both gradient and normal operations. • Thermo block made of silver. • Each block should accommodate 16 wells and having the ability to set up PCR with a specific temperature differential up to at least 5⁰ C across individual block. • System should have <ul style="list-style-type: none"> $\frac{3}{4}$ Temperature accuracy: ± 0.250 C (35⁰ C-99.9⁰ C) $\frac{3}{4}$ Temperature range: 4⁰ C-99⁰ C $\frac{3}{4}$ Temperature uniformity: <0.5⁰ C • System Should have on board Tm calculator facility to approximate the optimal annealing temperature. • Should be capable of testing temperatures at Denaturation, Annealing & Extension steps. • System should support PCR volumes ranging from 10-30 microlitres. • Memory: choice of saving up to 800 programs or unlimited to the memory stick. • System should have a USB port to transfer programs from machine. • Display: graphical user interface programming & colour touch screen. • Programmable heat lid cover from 50⁰ C to 105⁰ C. • Power: 220V/50~60Hz compatible. • Online UPS for back up should be provided along with. • Machine should be duly certified/ authorized for PCR process • Accessories: Necessary optimized reagents and plastic ware for standard and fast thermal cycling reactions to be provided. 	01 Nos.

Note:- The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the Research Section, AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.

[On the letterhead of firm]

ANNEXURE "2"
PRICE BIDFORM

To,

Dean (Research),
Research Section,
AIIMS, Jodhpur.

Dear Sir,

1. I/We Submitted the quotation for Enquiry No. "QUOTATION FOR SUPPLY OF THERMAL CYCLER AT RESEACH SECTION, AIIMS AGAINST THE INQUIRY NO. AIIMS/RES(06)/2019/383" DUE ON 23.09.2019 05.00 PM for Supply of Thermal Cyclers at Research Section, AIIMS Jodhpur".

2. I/We thoroughly examined, understood and accepted terms & conditions given in the enquiry document, failing which my quotation will be rejected out rightly.

3. I/We hereby offer to supply at the following rates.

S. No	Particular	Qty	Make	Price/Unit Exclusive of GST (INR)	GST/ Other Taxes	Price/ Unit Inclusive of GST (INR)	MRP
1.	Thermal Cyclers	01 Nos.					

Note:-

- The Bidder must quote only single Make & Model.**
- The Bidder must submit the GSTIN Registration and PAN Card self-attested copy with the quotation otherwise quotation will be REJECTED.**
- The bidder must quote their quotation only in above said format on the letter of firm otherwise quotation will be REJECTED.
- Catalog must be attached with quotation for technical evaluation.
- The Bidder must submit the sample of quoted make for technical evaluation at Research Section, Medical College, First Floor, AIIMS Jodhpur at the time of submission of their quotation. Failure to submit the sample at the time of submission of their quotation/offer will be summarily rejected.**

Date _____

(Name) _____

Place _____

Name of Firm/Company/Agency _____

GSTIN No.: _____

Bank Name:- _____

Bank Account No.: _____

IFSC Code:- _____

Branch Name: _____

Phone No. _____

Email: _____

(Signature of Authorized Person) _____

Seal: _____