



All India Institute of Medical Sciences,
Jodhpur

Dated: 23rd July, 2019

CORRIGENDUM

Invitation of quotation

For

Disposable Paediatrics Heated Wire Ventilator

Circuit

At

All India Institute of Medical Sciences, Jodhpur

Inquiry No. : Admin/Gen/32-06/2019-AIIMS.JDH

Inquiry Issue Date : 22nd July, 2019

Last Date of Submission : 29th July, 2019 at 05:00 PM..

The following revisions / modifications is made –**Annexure 1**

S.No.	Item Name	Specification	Qty.
1.	Disposable Paediatrics Heated Wire Ventilator Circuit	<p>1. Circuit should be in 1.6 meter in length and 15 mm diameter and should be compatible for both preterm neonates and children.</p> <p>2. It should be dual heated wire circuit i.e. both inspiratory and expiratory limbs should have heated wire.</p> <p>3. There should be a humidifier chamber along with this ventilator circuit.</p> <p>4. The ventilator circuit should be compatible with ventilators like drager babylog 8000 plus, SLE 5000, febian-accutronics and maquet servo-i. The appropriate connecter required for using this ventilator circuit with the above mentioned ventilators, should be available with the main circuit itself and no extra connectors form outside should be required.</p> <p>5. The circuit should be sterile in a proper packaging material.</p>	70 Nos.

Note:- The Bidder must submit the sample of quoted make for technical evaluation at Administrative Office, Medical College, Second Floor, AIIMS Jodhpur at the time of submission of their quotation. Failure to submit the sample at the time of submission of their quotation/offer will be summarily rejected.