

Invitation of quotation

For

Radiant Warmer

At

All India Institute of Medical Sciences, Jodhpur

Inquiry No.: : Admin/Gen/32-22(ii)/2019-AIIMS.JDH

Inquiry Issue Date : 29th January, 2020

Last Date of Submission : 04th February, 2020 at 05:00 PM.



All India Institute of Medical Sciences, Jodhpur

Basni Phase - II, Jodhpur – 342005, Rajasthan

Telefax: 0291- 2740741, email: procurement@aiimsjodhpur.edu.in

www.aiimsjodhpur.edu.in

Invitation of quotation for Radiant Warmer at AIIMS Jodhpur

Sealed Quotations are hereby invited by the undersigned on behalf of the Director, AIIMS Jodhpur for Radiant Warmer for the Institute as per terms & conditions mentioned below. The filled quotations along with all the required document must reach in the office of the undersigned on or before 04.02.2020 05.00 PM. The Envelope containing the quotation would please be sealed and super scribed as under:-

“QUOTATION FOR RADIANT WARMER AGAINST INQUIRY NO. ADMN/GEN/32-22(ii)/2019-AIIMS.JDH” DUE ON 04.02.2020 05.00 PM”

1. Terms & Conditions:

- A) The quotations received after this deadline & unsealed shall not be entertained under any circumstances whatsoever. In case of postal delay this Institute will not be responsible. **The offer Submitted Fax/Email shall not be considered and no correspondence will be entertained in this matter.**
- B) Quotations must be in the enclosed prescribed Performa on the letter head of the firm duly signed by the Proprietor/ Partner/ Director or their authorized representative, In case of signing of quotation by the authorized representative letter of authorization must be attached with the quotation. Quotation must be dropped in “**Quotation Box**” located in Administration Block of AIIMS, Jodhpur.
- C) Rates must be quoted in **Indian rupees** and as per the format specified taxes extra if any must be written separately.
- D) Rates must be quoted FOR basis (including Freight charges, Insurance, installation etc.)
- E) No overwriting or cutting is permitted in the rate. If found, the quotation shall be summarily rejected.
- F) The rates quoted must be valid for 60 days minimum from the date of opening of the quotation and silence of any tendered on this issue shall be treated as agreed with this condition.
- G) Becoming L1 will not be the criteria for awarding of purchase order unless the rates are reasonable & justified.
- H) RTGS/NEFT details need to be furnished by the supplier with the quotation on the letter head of supplier/firm/agency.
- I) The firm/agency may satisfy the following conditions and attach self-attested copy of the same with the quotation:
- Firm shall be registered with the Government of Rajasthan / Central Government.
 - The firm shall have valid GST/Other taxes and IT PAN.
 - **The firm should not be black listed by any Govt. Agency/Dept.**

INQUIRY NO. Admin/Gen/32-22(ii)/2019-AIIMS.JDH

- J) Quotations qualified by such vague and indefinite expressions such as “subject to prior confirmation”, “subject to immediate acceptance” etc. will be treated as vague offers and rejected accordingly. Any conditional quotation shall be rejected summarily.
- K) **Delivery Period** – within 30 days from Purchase order.
- L) **Liquidated Damage:** - If the supplier fails to deliver the material on or before the stipulated date, then a penalty at the rate of 0.5 % per week of the total order value shall be levied subject to maximum of 10% of the total order value.
- M) **Payment Terms:** Payment will be only after satisfactorily delivery / commissioning of material and after inspection by the AIIMS Jodhpur.
- N) **Disputes:** -In the event of any dispute or disagreement arising between the contractors and any other department of AIIMS Jodhpur with regards to the interpretation of “Terms & Conditions” of this inquiry, the same shall be referred to the Director, AIIMS Jodhpur whose decision will be final and binding upon the contractor.
- O) AIIMS, Jodhpur reserves the right to increase or decrease quantity and / or amount of work. Decision of Quantity of material in the AIIMS, Jodhpur will be final in this regard.
- P) AIIMS, Jodhpur reserves the right to reject any quotation or part or the whole of inviting quotation process without assigning any reason. Decision of the AIIMS, Jodhpur will be final in this regard.

2. Special Terms & Conditions:

- A) **Bidder must quote the product as per specification provided in Annexure 1.**
- B) **Catalog must be attached with quotation for technical evaluation.**
- C) **The Bidder must submit the GSTIN Registration and PAN Card self-attested copy with the quotation.**
- D) **The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.**

Administrative Officer

Encl.: Annexure 1 (Specification)
Annexure 2 (Format of price bid)

Annexure 1

Sr No.	Item Name	Qty.
1.	Radiant Warmer	04 Nos.

Specification of Radiant Warmer:-

1. Essential parts

- a) Quartz/ceramic/calrod based warming system with microprocessor based controls, probes & alarms
- b) Cart & bassinet
- c) Examination light
- d) X-ray cassette holder

2. Cart

- a) Should swivel on 4 wheels of at least 4" diameter with foot operated brakes on at least 2 front wheels

3. Dimensions

- a) Height - 180- 200 cms
- b) Width - 65-80 cms
- c) Depth - 100-120 cms
- d) Working level - 95-105 cm

4. Bassinet

- a) Collapsible transparent acrylic sidewalls

5. Mattress

- a) Width - 70 - 80 cms
- b) Length - 90 - 100 cms
- c) Thickness - 3-5 cm
- d) Material
 - i. Soft and easy to clean
 - ii. X-ray transparent
 - iii. Fire retardant
 - iv. Allows air to pass through but does not allow water to seep in

6. Bassinet tilt

- a) Should allow tilt for Trendelenburg as well as reverse Trendelenburg position

7. Warming system

- a) Quartz based heating system encapsulated heater tube placed in a parabolic reflector
- b) Should have single element quartz infrared heater
- c) Control - Microprocessor controlled
- d) Self test function performed at power on
- e) Digital display should show following parameters
 - i. Set temperature
 - ii. Present temperature of the baby
 - iii. Heater output
 - iv. The temperatures should be in numerical LED display
- f) Mode - Manual & skin (servo)

INQUIRY NO. Admin/Gen/32-22(ii)/2019-AIIMS.JDH

- g) Manual mode
 - i. Adjustable in steps from 0 to 100% in increments of 10%
 - ii. Heater power should be reduced to 50 - 60% after 10-15 minutes in manual mode for baby safety
- h) Skin mode (servo mode)
 - i. Set point range - 32 – 38 degrees C
- i) Skin temperature display
 - i. Accuracy - ± 0.2 degrees C
 - ii. Resolution - 0.1 degree C
- j) Temperature probe- Wire should be easy to clean and long lasting
- k) No need of temperature probe calibration

8. Alarms

- a) Audiovisual alarms with a display of text messages about the alarms
 - i. Probe failure
 - ii. Heater failure
 - iii. High and low baby temperature (more than 0.5 deg C difference)
 - iv. Power failure
 - v. System failure
 - vi. Silence/reset switch
 - vii. Should have Automatic cut-off of heater at 39°C for Over Temperature Protection

9. Heating unit - Should be swivable for accommodating X-Ray unit and should have self-lock facility.

10. Examination light

- a) Illuminance - at least 75 foot candles at mattress centre

11. Apgar Timer

- a) Timer with stopwatch facility
- b) Reset facility

12. I.V. stand

a) Strong IV stand (S.S) with height adjustable and facility to fix large number of infusion pumps

1. Monitor shelves - Two in number

2. Should have two drawers

15. X- Ray cassette holder

- a) Sliding holder located just below undersurface of bassinet, with markings to help placement of cassette

16. Power consumption - Less than 1 K.W

17. All metal parts of the equipment should be corrosion resistant and Epoxy/Powder coated

18. All consumables required for installation and standardization of system to be given

INQUIRY NO. Admin/Gen/32-22(ii)/2019-AIIMS.JDH

free of cost

19. Standards, safety and training

- a) Should be European CE approved product and certificate needs to be submitted along with the catalogue
- b) Manufacturer should be ISO 13485 certified for quality standards
- c) Electrical safety conforms to standards for electrical safety IEC 60601-1 (OR EQUIVALENT international/national standard)General requirement for Electrical safety of Medical Equipment
- d) Should have type of protection against electric shock: Class 1
- e) Should have degree of protection against electric shock: Type B

20. Power supply and electrical specifications

- a) Power input to be 220-240VAC, 50Hz
 - b) Suitable Autovoltage corrector with spike protector should be available
 - c) Should have overload protection at 4A
 - d) Should have leakage current <0.5mA
 - e) Should have earth resistance <0.2 Ohm
 - f) Should be Low Noise
 - g) Fuse Rating should be
 - i. For main supply: for 230V: 4A & for 110V: 8A
 - ii. For examination lamp: for 230V: 2A & for 110V: 2A

21. Environmental factors

- a) Shall meet IEC-60601-1-2 :2001 General Requirements of Safety for Electromagnetic Compatibility or should comply with 89/366/EEC; EMC directive
- b) The unit shall be capable of being stored continuously in ambient temperature of 10-55 deg C and relative humidity of 15-90%
- c) The unit shall be capable of operating continuously in ambient temperature of 20-40 deg C and relative humidity of 15-90%

22. Documentation

- a) User/Technical/Maintenance manuals to be supplied in English
- b) List of important spares and accessories with their part number and costing

23. Essential accessories to be supplied at initial purchase with each piece of equipment

- a) Reusable temperature probes (full set): 10 nos/per equipment.

24. The rates of consumable accessories should also be quoted separately

25. Onsite physical demonstration/training of the equipment to all the end users with all the requested facilities will be

Note:- The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.

[On the letterhead of firm]

ANNEXURE "2"
PRICE BIDFORM

To,

Administrative Officer,
AIIMS, Jodhpur.

Dear Sir,

1. I/We Submitted the quotation for Enquiry No. "QUOTATION FOR RADIANT WARMER AT AIIMS AGAINST THE INQUIRY NO. Admn/Gen/32-22(ii)/2019-AIIMS.JDH" DUE ON 04.02.2020 05.00 PM for Radiant Warmer at AIIMS Jodhpur".
2. I/We thoroughly examined, understood and accepted terms & conditions given in the enquiry document, failing which my quotation will be rejected out rightly.
3. I/We hereby offer to supply at the following rates.

S. No	Particular	Qty	Required Make	Price/Unit Exclusive of GST (INR)	GST/ Other Taxes	Price/ Unit Inclusive of GST (INR)	MRP
1.	Radiant Warmer	04 Nos.					

Note:-

1. The Bidder must quote only recommended Make & Model.
2. The Bidder must submit the GSTIN Registration and PAN Card self-attested copy with the quotation otherwise quotation will be **REJECTED**.
3. The bidder must quoted their quotation only in above said format on the letter of firm otherwise quotation will be **REJECTED**.
4. Catalog must be attached with quotation for technical evaluation.
5. The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.

Date _____

(Name) _____

Place _____

Name of Firm/Company/Agency _____

GSTIN No.: _____

Bank Name:- _____

Bank Account No.: _____

IFSC Code:- _____

Branch Name: _____

Phone No. _____

Email: _____

(Signature of Authorized Person) _____

Seal: _____