



Date: - 05th April, 2017

Corrigendum
for
tender
for
Low Temperature Hydrogen Peroxide Plasma Gas
Sterilizer for the Department of CSSD

NIT Issue Date	: 16 th February, 2017
NIT No.	: Admn/Tender/111/2016-AIIMS.JDH
Pre-Bid Meeting	: 01 st March, 2017 at 03:00 PM
Earlier Last Date of Submission	: 10 th April, 2017 at 03:00 PM
Extended Last Date of Submission	: 17 th April, 2017 at 03:00 PM
Bid opening	: 18 th April, 2017 at 03:00 P.M

The following revised and additional specification will be added:-

1. Page No. 11, In technical specification, Point No. 05:

For

Should have sterilization chamber more than 100 Liters usable volume with removable shelf.

Read

Should have a sterilization chamber of usable volume of more than 100 liters or more with removable shelves.

2. Page No. 11, In technical specification, after Point No. 23:

Add Point No. 24:

Sterilant cassette/ bottle should be such that it can be stored at room temperature.

3. Page No. 11, In technical specification, after Point No. 24:

Add Point No. 25:

Sterilizer should be able to detect excess/residual moisture on Instruments

4. Page No. 11, In technical specification, after Point No. 25:

Add Point No. 26:

It is preferable that the sterilizer should be recommended by leading laparoscope instrument manufacturers like Stryker, Karl Storz, Olympus etc for safety of telescopes and cameras.

**Administrative Officer
AIIMS, Jodhpur**