



Date: - 22nd July, 2020

Corrigendum-2
for
Full HD Video Endoscopy System (Therapeutic) for
the Department of Surgical Gastroenterology

NIT Issue Date	: 20 th May, 2020
NIT No.	: Admn/Tender/12/2020-AIIMS.JDH
Pre-Bid Meeting	: 03 rd June, 2020 at 03:45 PM
Earlier Last Date of Submission	: 30 th July, 2020 at 03:00 PM
Extended Last Date of Submission	: 05 th August, 2020 at 03:00 PM
Bid opening	: 06 th August, 2020 at 03:15 P.M

The following revised and additional specification will be added:-

1. In Corrigendum, Page No. 1, Point No. 2:

For

Should have real time optical/digital chrome endoscopy imaging such as NBI/Optical enhancement or equivalent. Should have dual/close focus capacity for providing detailed mucosal study

Read as:

Should have real time optical chrome endoscopy imaging such as NBI/ BLI/ Optical enhancement or equivalent. Should have dual/close focus capacity for providing detailed mucosal study

2. In Corrigendum, Page No. 1, Point No. 4:

For

Depth of field: 4 to 60 mm or better

Read as:

Depth of field: 4/5 to 60 mm or better

3. In Corrigendum, Page No. 1, Point No. 5:

For

Should be compatible with real time optical /digital chrome endoscopy imaging such as NBI/optical enhancement or equivalent

Read As:

Should be compatible with real time optical chrome endoscopy imaging such as NBI / BLI / Optical enhancement or equivalent. Should have Picture-in-Picture (PIP) / Picture-out-Picture (POP) function.

4. In tender, Page No. 10, Point No. 3

For

Xenon short arc Ozone free 300 Watt lamp

Read As

Xenon short arc Ozone free 300 Watt lamp or Multi LED's (3 or more LED) equivalent to 300W Xenon lamp intensity

5. In Corrigendum, Page No. 2, Point No. 7:**For**

Real time optical chrome endoscopy imaging such as NBI/Optical enhancement or equivalent capability. 300W xenon light source with halogen back up. Two xenon spare lamps to be supplied along with the machine.

Read As

Real time optical chrome endoscopy imaging such as NBI /BLI / Optical enhancement or equivalent. 300W xenon light source with halogen back up. One xenon spare lamps to be supplied along with the machine.

6. Addition of Points :

- 1) Original make trolley with scope and monitor holder mounted on the trolley should be supplied.
- 2) Cost of additional Xenon bulb to be quoted separately and that will freeze for next 2 years
- 3) All the equipment's and ancillary items must be European CE (Four digit notified body number) / US FDA / Bureau of Indian Standard (BIS) approved.