



Date: - 13th February 2019

Corrigendum

For

Height Adjustable Grossing Station with Inbuilt exhaust with Inbuilt Enclosed Formalin Tanks & Optional for the Department of Pathology

NIT Issue Date	: 15 th November, 2018
NIT No.	: Admn/Tender/137/2018-AIIMS.JDH
Pre-Bid Meeting	: 26 th November, 2018 at 03:00 PM
Earlier Last Date of Submission	: 14 th February, 2019 at 03:00 PM
Extended Last Date of Submission	: 28 th February, 2019 at 03:00 PM
Bid opening	: 01 st March, 2019 at 03:15 P.M

The following revised and additional specification will be added:-

1. Page No. 4, Terms & Conditions Point No. 9 (ii)

For

Delivery and Installation:

All the goods ordered shall be delivered and Installed at AIIMS, Jodhpur within 90 days from the date of opening of Letter of Credit for shipment.

Read

Delivery and Installation: For goods imported directly from abroad:

All the goods ordered shall be delivered at AIIMS, Jodhpur within 150 days from the date of opening of Error Less (without any further need of amendment in LC document) Letter of Credit for shipment.

2. Page No. 10, Annexure I, S No. 1, in technical specification, point No. 1

For

The equipment should be a floor mounted model.

Read

The equipment should be a floor mounted Height adjustable model.

3. Page No. 10, Annexure I, S No. 1, in technical specification, point No. 2

For

The station should be made of non-corrosive high grade stainless steel

Read

The station should be made of non-corrosive high 304 grade stainless steel

4. Page No. 10, Annexure I, S No. 1, in technical specification, point No. 7

For

Sink should have removable filters on drain to trap debris/tissue bits

Read

Should have mixing faucet for both Hot and Cold Water

5. Page No. 10, Annexure I, S No. 1, in technical specification, point No. 14**For**

Top mounted fluorescent light fixture

Read

Should have Top mounted LED light fixture

6. Page No. 10, Annexure I, S No. 1, in technical specification, point No. 15**For**

Halogen light mounted on a flexible arm

Read

Should have Fluorescent Magnifier Light on flexible spring balanced arm

7. Page No. 10, Annexure I, S No. 1, in technical specification, point No. 18**For**

Polyethylene dissecting board

Read

Polypropylene (Blue board)

8. Page No. 10, Annexure I, S No. 1, in technical specification, point No. 23**For**

There should be facility for video, audio recording and photography attachment. The photography attachment should have facility for enlargement.

Read

Should have Camera Mount Arm which securely holds camera, adjustable ball and socket for ease of adjustment and better coverage.

Should be supplied with Digital SLR Camera (canon/Nikon) with 18-55 lens, CMOS Sensor, 16 GB card for recording good quality of photographs with computer interphase (18.0 Mega Pixel) with HDMI cable.

9. Page No. 10, Annexure I, S No. 1, in technical specification, after point No. 24.**Add point No. 25**

Should have Cassette Holders which conveniently holds cassettes within arm's reach and removable boxes mount to rail on front side, set of three boxes

10. Page No. 10, Annexure I, S No. 1, in technical specification, after point No. 25.**Add point No. 26**

Should have Scale – Pole & Bracket attached on side of the Grossing Station consisting of Digital Autopsy Scale of 13.6 Kg x 0.1 Kg capacity along with bow, ring and pan for weighing organs and specimens.

11. Page No. 10, Annexure I, S No. 1, in technical specification, after point No. 26**Add points No. 27**

Height (Fully Extended): 96 in./ 243 cm Width: 30 in./ 76 cm

Table Dimension:

Length: 48 in./ 121 cm

Height (Lowest): 84 in./ 213 cm