



Date: - 18<sup>th</sup> April, 2019

Corrigendum  
For  
Operating Microscope for Spine Surgery for the  
Department of Orthopaedics

NIT Issue Date	: 1 <sup>st</sup> December, 2018
NIT No.	: Admn/Tender/143/2018-AIIMS.JDH
Pre-Bid Meeting	: 10 <sup>th</sup> December, 2018 at 05:00 PM
Earlier Date of Submission	: 15 <sup>th</sup> April, 2019 at 03:00 PM
Extended Date of Submission	: 15 <sup>th</sup> May, 2019 at 03:00 PM
Bid opening	: 16 <sup>th</sup> May, 2019 at 03:15 P.M

The following revised and additional specification will be added:-

1. Page No. 1 of 1<sup>st</sup> Corrigendum, Point 3<sup>rd</sup>:

**For**

Preferably Advanced digital Completely Integrated 3CCD HD video camera system suitable for connection to PC, color monitor. Video speed focus.

**Read**

Advanced digital Completely Integrated 3CCD HD video camera system suitable for connection to PC, color monitor. The camera should not be visible from outside, it should be integrated in the microscope head, preferably video speed focus.