



Date: - 13th February 2019

Corrigendum
For
Operating Microscope for Spine Surgery for the
Department of Orthopaedics

NIT Issue Date	: 1 st December, 2018
NIT No.	: Admn/Tender/143/2018-AIIMS.JDH
Pre-Bid Meeting	: 10 th December, 2018 at 05:00 PM
Earlier Last Date of Submission	: 14 th February, 2019 at 03:00 PM
Extended Last Date of Submission	: 28 th February, 2019 at 03:00 PM
Bid opening	: 01 st March, 2019 at 03:15 PM

The following revised and additional specification will be added:-

- 1. Page No. 10, Heading: “AutoDrape”:**
For
System should be capable of using an auto drape system by push of a button
Read
Preferably System should be capable of Draping the microscope by push of a button.
- 2. Page No. 10, Heading: “Floor Stand” , Point No. 4 :**
For
System should have facility for Auto Air evacuation system for drapes(Auto drapes).
System should have overhead display for showing important parameters
Read:
System should Preferably have facility for Auto Air evacuation system for drapes(Auto drapes). System should have overhead display for showing important parameters.
- 3. Page No. 10, Heading: “INTEGRATED DIGITAL VIDEO CAMERA SYSTEM” :**
For
Advanced digital Completely Integrated 3CCD HD video camera system suitable for connection to PC, color monitor. Video speed focus.
Read
Preferably Advanced digital Completely Integrated 3CCD HD video camera system suitable for connection to PC, color monitor. Video speed focus.