



Date: - 25th January 2021

Corrigendum-2
for
Biopolar TRUP and TUEB Instruments for the
Department of Urology

NIT Issue Date	: 26 th November, 2020
NIT No.	: Admn/Tender/16-2/2020-AIIMS.JDH
Pre-Bid Meeting	: 07 th December, 2020 at 04:00 PM
Earlier Last Date of Submission	: 28 th January, 2020 at 03:00 PM
Extended Last Date of Submission	: 09 th February, 2021 at 03:00 PM
Bid opening	: 10 th February, 2021 at 03:15 P.M

The following revised and additional specification will be added: -

- Page No. 1, In technical specification, After Toggle Button No. 2:**
For
Telescope, 0 Degree, Diameter 4 mm, Length 30 cm, Enlarged view, Autoclavable, Fiber-optic light transmission incorporated 1 Nos.
Read
Deleted
- Page No. 1, In technical specification, After Toggle Button No. 5:**
For
Two way Irrigation Port for 24 Fr/22Fr sheath for continuous flow TUR 2 Nos.
Read
Deleted
- Page No. 1, In technical specification, After Toggle Button No. 6:**
For
Bipolar Passive Working element with LLS (Logical Locking Snap -FIT) – . Motion by mean of spring movable thumb ring, in rest position the electrode is inside the sheath. Qty. 2
Read
Bipolar Passive Working element – Motion by mean of spring movable thumb ring, in rest position the electrode is inside the sheath Qty. 2
- Page No. 1, In technical specification, After Toggle Button No. 11:**
For
Cystoscope sheath 19.5 Fr. 1 Nos.
Read
Cystoscope sheath 19-20 Fr. 1 Nos.
- Page No. 2, In Accessories , After Toggle Button No. 3:**
For
Bipolar cutting loop electrode, compatible with 26 fr resectoscope, 0.35 mm wire diameter 15 Nos.

Read

Bipolar cutting loop electrode, compatible with 26 fr resectoscope, 0.35 mm wire diameter 12
Nos.

6. Page No. 2, In Accessories, After Toggle Button No. 4:**For**

Bipolar vaporization mushroom shaped electrode 8 Nos.

Read

Bipolar Vapo-Enucleation (Mushroom Shape) electrode 6 Nos.

7. Page No. 2, In Accessories, After Toggle Button No. 7:**For**

Bipolar Needle Electrode small 2 Nos.

Read

Bipolar Needle Electrode 2 Nos.

8. Page No. 2, In Accessories, After Toggle Button No. 8:**For**

Bipolar loop Electrode for plasma vaporization small ,medium & large 4 Pcs. each

Read

Bipolar loop Electrode for plasma vaporization 6 Pcs.

9. Page No. 2, In Accessories, After Toggle Button No. 11:**For**

Rotatable Irrigation Port, 2 Stopcocks 2 Nos.

Read

Deleted

10. Page No. 2, In Accessories, Toggle Button No. 15:**For**

Tray for sterilization and storage of loops and electrodes from reputed make. 5Nos.

Read

Tray for sterilization and storage of Telescope, loops and electrodes from reputed make.

1 for each telescope, 1 for resectoscope, working element and other endoscopic instruments and 1 for loops and electrodes.

11. Page No. 2, In Accessories, Toggle Button No. 16:**For**

All Instruments should be USFDA &/or European four digit CE certified & from same maker and should have direct service center

Read

- All Instruments should be USFDA &/or European four digit CE certified & from same maker.
- Bipolar resection set should be compatible with the Electrode Surgical Generator from reputed make installed in our institution. (Erbe, Olympus, Covidien/ Medtronic).