



Date: -14<sup>th</sup> July, 2017

Corrigendum  
For  
Tender for  
Multiparamonitor for the Department of Pediatric  
Surgery

NIT Issue Date	: 03 <sup>rd</sup> April, 2017
NIT No.	: Admn/Tender/34/2017-AIIMS.JDH
Pre-Bid Meeting	: 17 <sup>th</sup> April, 2017 at 03:45 PM
Earlier Last Date of Submission	: 18 <sup>th</sup> July, 2017 at 03:00 PM
Extended Last Date of Submission	: 02 <sup>nd</sup> August, 2017 at 03:00 PM
Bid opening	: 03 <sup>rd</sup> August, 2017 at 03:30 P.M

**The following revised and additional specification will be added:-**

- 1. Page No. 11, In Technical Specification, Point No. 2:**  
**For**  
Should be US FDA and European CE certified.  
**Read:**  
It should be US FDA or European CE certified.
- 2. Page No. 11, In Technical Specification, Point No. 7:**  
**For**  
Medical grade, TFT Flat screen, slim size, at least 10 inch display.  
**Read:**  
Medical grade, TFT Flat screen, slim size, at least 12 inch display.
- 3. Page No. 12, In Technical Specification, Point No. 17, Noninvasive blood pressure, After Point No. 17.8:**  
Added Point No: 17.9:  
It should also give perfusion index in range of 0-20%.
- 4. Page No. 12, In Technical Specification, Point No. 20:**  
**For**  
Point No. 20.  
Memory storage at least 24 hours.  
**Read:**  
Memory storage at least 72 hours.

**5. Page No. 12, In Technical Specification, After Point No. 30:**

Added Point No. 31:

It should have ST mapping i.e. continuous ST segment monitoring by incorporating graphical display for the ST segment changes.

**6. Page No. 12, In Technical Specification, After Point No. 31:**

Added Point No. 32:

It should have QT analysis to identify problems with ventricular functioning.