



Date: - 27th November, 2019

Corrigendum-2
for
4K Laparoscopy Imaging System with ICG
compatibility for the Department of Surgical
Gastroenterology

NIT Issue Date	: 21 st August, 2019
NIT No.	: Admn/Tender/39/2019-AIIMS.JDH
Pre-Bid Meeting	: 02 nd September, 2019 at 03:00 PM
Earlier Last Date of Submission	: 28 th November, 2019 at 03:00 PM
Extended Last Date of Submission	: 16 th December, 2019 at 03:00 PM
Bid opening	: 17 th December, 2019 at 03:15 P.M

The following revised and additional specification will be added:-

1. In Corrigendum, Page no. 2, Point no. 2 , Point a)

For

A Powerful 300 Watt Xenon Lamp or equivalent LED light source.

Read as:

Two powerful 300 Watt or more Xenon or equivalent LED light source to provide 4K and ICG output for independent working.

2. In Corrigendum, Page no. 2, Point no. 2 , Point e)

For

300W Xenon light source with spare lamp (10 extra bulbs) having " NBI technology" / ICG technology

Read as:

300W Xenon light source with 5 spare bulbs having "NBI technology" / ICG technology

3. In Corrigendum, Page no. 5, Point no. 8 , Point (b), "Image and Video recording and data archiving management system":

For

Should be a 4K UHD / full HD medical grade video and still images recorder

Read as:

Should be able to record 4K High definition medical grade video and still images recorder

4. In Technical specification, Page No. 12, Point No. 8, sub-point No. (f)"

For

Should have dual channel simultaneous recording for 4K UHD /full HD

Read as:

Should have dual channel simultaneous recording for 4K High Definition images and videos.

5. In Corrigendum, Page no. 5, Point no. 8 , Sub-Point (O), “Image and Video recording and data archiving management system”:

For

Should have support for physician print, media, annotations, patients field, procedure Settings and individual surgeon profile & preferences.

Read as: Optional

6. In Corrigendum, Page no. 5, Point no. 8 , Sub-Point (r), “Image and Video recording and data archiving management system”:

For

Following accessories should be supplied: Cordless mouse and cordless silicone keypad and foot pedal control

Read as: Optional; provide if available.

7. In Corrigendum, Page no. 6, Point no. 9:

For

All products should be same manufacturer & should be US FDA or European CE approved (with four digit notified body number)

Read As

All products should be from same manufacturer & must be US FDA or European CE approved (with four digit notified body number).