



Date: - 17th December, 2019

Corrigendum
for
Mobile exercise test system with ergometer control for
the Department of Physiology

NIT Issue Date	: 04 th November, 2019
NIT No.	: Admn/Tender/55/2019-AIIMS.JDH
Pre-Bid Meeting	: 14 th November, 2019 at 03:00 PM
Earlier Last Date of Submission	: 16 th December, 2019 at 03:00 PM
Extended Last Date of Submission	: 30 th December, 2019 at 03:00 PM
Bid opening	: 31 st December, 2019 at 03:15 P.M

The following revised and additional specification will be added:-

1. Page No. 10, Technical Specification, Point No. 19:

For

Provide digital output of data recorded to be used to synchronize with other recorded physiological parameters on other acquisition systems

Read as:

Provide digital output of raw data recorded at least in CSV or any other general format to be used for further in-depth higher order analysis.