



Date: - 09th January, 2020

Corrigendum
for
Transcutaneous Bilirubinometer for the Department of
Pediatric Surgery

NIT Issue Date	: 04 th November, 2019
NIT No.	: Admn/Tender/59-2/2018-AIIMS.JDH
Pre-Bid Meeting	: 14 th November, 2019 at 03:45 PM
Earlier Last Date of Submission	: 15 th January, 2020 at 03:00 PM
Extended Last Date of Submission	: 30 th January, 2020 at 03:00 PM
Bid opening	: 31 st January, 2020 at 03:15 P.M

The following revised and additional specification will be added:-

- 1. Page No. 10, Point 8:**
For: -
Light source should be Pulse xenon arc lamp
Read: -
Light source should be pulse xenon arc lamp or LED lamp.
- 2. Page No. 10, Point 11 -**
For
Should have detectors with Silicon photodiodes
Read
Should have silicon photodiodes
- 3. Page No. 10, Point 17 -**
For
It should measure at least 400 single measurements when fully charged.
Read
It should measure at least 100 single measurements when fully charged.