



Date: - 18th February, 2020

Corrigendum
for
Video Laryngoscope with C-MAC Monitors for
Modular Operation Theater

NIT Issue Date	: 31 st December, 2019
NIT No.	: Admn/Tender/60/2019-AIIMS.JDH
Pre-Bid Meeting	: 13 th January, 2020 at 03:00 PM
Earlier Last Date of Submission	: 17 th February, 2020 at 03:00 PM
Extended Last Date of Submission	: 02 nd March, 2020 at 03:00 PM
Bid opening	: 03 rd March, 2020 at 03:15 P.M

The following revised and additional specification will be added:-

**1. Page No. 10, In technical specification, C Mac Monitor, Bullet No. 02:
For**

Video recording and still image recording facility should be available (Minimum 3-4 GB internal memory) on data card or USB with JPEG and MPEG4 format which can be easily transferred to the computer/laptop. Connecting cable should be provided

Read

Video recording and still image recording facility should be available (Minimum 3-4 GB internal memory) on data card or USB with JPEG and MPEG4 format which can be easily transferred to the computer/laptop. Connecting cable should be provided & Should have HDMI port with 2 video outputs.

**2. Page No. 10, In technical specification, C Mac Monitor, After Bullet No. 06:
Added bullet No. 7:**

- The Monitor, Scopes, Movable stand and all accessories should be supplied through single OEM.