



Date: - 01st February, 2017

Corrigendum
for
tender
for
Hemodialysis Machine for the Department of
Nephrology

NIT Issue Date	:	18 th November 2016
NIT No.	:	Admn/Tender/88/2016-AIIMS.JDH
Pre Bid Meeting held on	:	30 th November, 2016 at 03:00 PM
Earlier Date of Submission	:	07 th February, 2017 at 03:00 PM
Extended Last Date of Submission	:	14 th February, 2017 at 03:00 PM
Bid Opening	:	15 th February, 2017 at 03:45 PM

The following revised and additional specification will be added:-

- 1. Page No. 12, System Configuration Accessories, Spares and consumables, Toggle Button No. 2:**
For
Company must supply Bacterial filters-6 sets extra and 100 dialyzers with tubings free of cost.
Read
Company must supply total -6 Bacterial filters sets extra and 100 dialyzers with tubings free of cost.
- 2. Page No. 12, Power supply, Toggle Button No. 3:**
For
One 3 KVA online UPS with at least 30 min power backup must be supplied with each machine.
Read
One 3 KVA online UPS with at least 30 min power backup should be supplied with each machine(optional).
- 3. Page No. 12, In Documentation, after Toggle Button No. 10:**
Added Point No. 10:
Machine should have Closed System hydraulics -volumetric fluid balancing system and UF monitor.
- 4. Page No. 12, In Documentation, after Toggle Button No. 11:**
Added Point No. 12:
Monitoring of set alarm limits (That is automatic alarm limit set according to the blood pump rate delivery, UF rate and Venus chamber level adjustment).

Administrative Officer
AIIMS, Jodhpur