



Date: - 22nd September, 2020

Corrigendum
for
Rate Contract
For
Supply of Liquid Medical Oxygen

NIT No.	: Admn/RC/03/2020-AIIMS.JDH
NIT Issue Date	: 29 th August 2020
Pre Bid Meeting	: 09 th September 2020 at 03.00 PM
Earlier Last Date of Online Submission	: 21 st September 2020 at 03.00 PM
Extended Last Date of Online Submission	: 29 th September 2020 at 03.00 PM
Bid Opening Date	: 30 th September 2020 at 03.45 PM

The following revised and additional conditions will be added:-

1. Page No. 2, EMD Payment:

For

The bidder shall be required to submit the Earnest Money Deposit (EMD) for an amount of **Rs. 1,08,000/- (Rupees One Lakh Eight Thousand Only)** by way of demand drafts or Bank Guarantee only. The demand drafts shall be drawn in favour of "**All India Institute of Medical Sciences, Jodhpur**" payable at Jodhpur. **The demand drafts or Bank Guarantee for earnest money deposit must delivered to the AIIMS, Jodhpur on or before bid closing date/time (submitted only in Dispatch/Received section).** The EMD of the successful bidder shall be returned after the successful submission of Bank Guarantee/ Security Deposit and for unsuccessful bidder(s) it would be returned after award of the contract. Bid(s) received without demand drafts of EMD will be rejected.

Read as:

The bidder shall be required to submit the Earnest Money Deposit (EMD) for an amount of **Rs. 1,08,000/- (Rupees One Lakh Eight Thousand Only)** by way of demand drafts or Bank Guarantee only. The demand drafts shall be drawn in favour of "**All India Institute of Medical Sciences, Jodhpur**" payable at Jodhpur. **The demand drafts or Bank Guarantee for earnest money deposit must delivered to the AIIMS, Jodhpur within 07 days of last date/ time of Online submission (submitted only in Dispatch/Received section).** The EMD of the successful bidder shall be returned after the successful submission of Bank Guarantee/ Security Deposit and for unsuccessful bidder(s) it would be returned after award of the contract. Bid(s) received without demand drafts of EMD will be rejected.

2. **Page No. 5, General Term & Conditions, Point No. 6, Delivery:**

For

The items will have to be supplied at manifold room in AIIMS Jodhpur premises. No transportation/ cartage charges will be provided for the same. All the aspects of safe delivery shall be the exclusive responsibility of the supplier.

Read as:

The items will have to be supplied at manifold room in AIIMS Jodhpur premises. No transportation/ cartage charges will be provided for the same. All the aspects of safe delivery shall be the exclusive responsibility of the supplier. **The delivery of Liquid Medial Oxygen shall be done in the licensed LMO Storage Tanks in the premise of AIIMS, Jodhpur in 72-96 Hrs. from time of raising the indent by the concerned authority of Hospital.**

3. **Page No. 5, General Term & Conditions, Point No. 11, Payment clause, After Para 03:**

Added as Para 04:

Payment shall be released within 15 (fifteen) working days after the successful supply of the ordered goods against the submission of the Inspection certificate by the AIIMS, Jodhpur

4. **Page No. 9, Special Terms & Conditions, Point No. 2:**

For

Guarantee Certificate: - The gases supplied against this order shall be deemed to bear a warranty of the contractor against deterioration, defective material, workmanship and performance for a period of 24 months from the date of receipt of stores at AIIMS, Jodhpur. If during this period the stores supplied are found to be defective in performance or deteriorated, the contractor shall be responsible for all consequences.

Read as:

Deleted.

5. **Page No. 12, Annexure-III, Point No.3:**

For

Following specifications of LMO, vaporizer and regulator to be followed.

- a) Should be certified for medical use as per IP 2014.
- b) Should not contain less than 99.6% V/V of oxygen.
- c) Should not contain more than 5ppm carbon monoxide.
- d) Should not contain more than 30ppm carbon dioxide.
- e) Should be free of halogen. Should be free of moisture

Note: - Provide undertaking each time for compliance & each time LOT is filled.

Read as:

Following specifications of LMO to be followed.

- f) Should be certified for medical use as per IP 2014.
- g) Should not contain less than 99.6% V/V of oxygen.
- h) Should not contain more than 5ppm carbon monoxide.
- i) Should not contain more than 30ppm carbon dioxide.
- j) Should be free of halogen. Should be free of moisture

Note: - Provide undertaking each time for compliance & each time LOT is filled.