



Date: -08<sup>th</sup> June, 2017

Corrigendum  
For  
Rate Contract  
For  
Implantable Hearing Devices  
For the Department of ENT

NIT No.	: Admn/RC/01/2017-AIIMS.JDH
NIT Issue Date	: 01 <sup>st</sup> April, 2017
Pre Bid Meeting	: 13 <sup>th</sup> April, 2017 at 03:00 PM
Earlier Last Date of Submission	: 05 <sup>th</sup> June, 2017 at 03:00 PM
Extended Last Date of Submission	: 21 <sup>st</sup> June, 2017 at 03:00 PM
Bid Opening	: 22 <sup>nd</sup> June, 2017 at 03:00 PM

**The following revised and additional specification will be added:-**

- 1. Page No. 13 Point No. 5, Warranty, Sub Point No. b:**  
**For**  
External Components excluding cables, Speech Processor, battery packs and coil: 3 years from the date of switch on.  
**Read**  
External Components including Speech Processor and excluding cables, battery packs and coil: 3 years from the date of switch on.
- 2. Page No. 13 Point No. 5, Warranty, Sub Point No. c:**  
**For**  
Consumables such as batteries chargers, rechargeable batteries – 2 years from the date of switch on.  
**Read**  
Consumables such as batteries chargers, rechargeable batteries – 1 years from the date of switch on.
- 3. Page No. 13 Point No. 5, External components (not covered by warranty), Sub Point No. a.:**  
**For**  
Robust and long lasting. Should be Sweat / Splash resistant.  
**Read**  
Robust and long lasting. Resistance against Non-Condensing Moisture.

**4. Page No. 19, Point No. 02, Intra-Cochlear Electrode Array, Sub Point No.B:**

**For**

Availability of various options for active electrodes, when needed for use in cases of different cochlear anomalies, at the same price. (Kindly provide details in brochure and compliance sheet)

**Read**

The electrode array should be usable for patients with different cochlear anomalies (either the standard electrode array or multiple options – depending upon the availability with the manufacturer – at the same price)

**5. Page No. 19, Point No. 02, Speech Processor, Sub Point No. C:**

**For**

Should have Wide Input Dynamic Range of 75DB or more with Automatic Sound Management to adapt automatically to different challenging environments. The sound management should be fully automatic with no manual switches to change the program.

**Read**

Should have Wide Input Dynamic Range of 75DB or more with automatic sound management (which also includes gain and sensitivity controls) to adapt automatically to different challenging environments. The sound management should be fully automatic with no manual switches to change the program.

**6. Page No. 19, Point No. 4, In Power Source:**

**For**

Powered by rechargeable cells/Rechargeable battery pack. Recharging equipment & Rechargeable Zinc air cells (at least three or more)/battery pack to be supplied by the manufacturer. Commitment for their free replacement, if they malfunction, within warranty period.

**Read**

Powered by rechargeable cells/Rechargeable battery pack. Recharging equipment & Rechargeable cells/battery packs (three or more Sets) to be supplied by the manufacturer. Commitment for their free replacement, if they malfunction, within warranty period. Also provide 10 or more button cells for use in emergency.

**7. Page No. 19, Point No. 5, External Components, Sub Point No. B:**

**For**

Two sets of all spares including spare cables and coils and commitment towards free replacement of external components not covered by warranty to be provided for a minimum period of 3 years from date of implantation.

**Read**

Two sets of all spares including spare cables and commitment towards free replacement of external components not covered by warranty to be provided for a minimum period of 3 years from date of implantation.

**8. Page No. 20, Point No. 16, Warranty, Sub Point No. B:**

**For**

External Components, including Speech Processor & battery packs, cables & coils: 3 years from the date of switch on.

**Read**

External Components, including Speech Processor: 3 years from the date of switch on.

**9. Page No. 20, Point No. 16, Warranty, Sub Point No. C:**

**For**

Consumable such as batteries chargers, rechargeable batteries - 2 years from the date of switch on.

**Read**

Consumable such as cables, coils, batteries chargers, rechargeable batteries/battery packs and all others - 1 years from the date of switch on.

**10. Page No. 20, Point No. 16, Warranty, Sub Point No. D:**

**For**

Should quote CMC for two years after warranty.

**Read**

Deleted

**11. Page No. 20, Point No. 18:**

**For**

At least one year of speech and rehabilitative therapy, which would include mapping etc.

**Read**

At least one year of support for speech and rehabilitative therapy, which would include mapping support etc.

**12. Page No. 21, Point No. 03, Speech Processor, Sub Point No. C.**

**For**

Should have Wide Input Dynamic Range of 75DB or more with Automatic Sound Management to adapt automatically to different challenging environments. The sound management should be fully automatic with no manual switches to change the program.

**Read**

Should have Wide Input Dynamic Range of 75DB or more with automatic sound management (which also includes gain and sensitivity controls) to adapt automatically to different challenging environments. The sound management should be fully automatic with no manual switches to change the program.

**13. Page No. 21, Point No. 04, In Power Source:**

**For**

Powered by rechargeable cells/Rechargeable battery pack. Recharging equipment & Rechargeable Zinc air cells (at least three or more)/battery pack to be supplied by the manufacturer. Commitment for their free replacement, if they malfunction, within warranty period.

**Read**

Powered by rechargeable cells/Rechargeable battery pack. Recharging equipment & Rechargeable cells/battery packs (three or more Sets) to be supplied by the manufacturer. Commitment for their free replacement, if they malfunction, within warranty period. Also provide 10 or more button cells for use in emergency.

**14. Page No. 21, Point No. 5, External Components, Sub Point No. B:**

**For**

Two sets of all spares including spare cables and coils and commitment towards free replacement of external components not covered by warranty to be provided for a minimum period of 3 years from date of implantation.

**Read**

Two sets of all spares including spare cables and commitment towards free replacement of external components not covered by warranty to be provided for a minimum period of 3 years from date of implantation.

**15. Page No. 22, Point No. 16, Warranty, Sub Point No. B:**

**For**

External Components, including Speech Processor & battery packs, cables & coils: 3 years from the date of switch on.

**Read**

External Components, including Speech Processor: 3 years from the date of switch on.

- 16. Page No. 22, Point No. 16, Warranty, Sub Point No. C:**  
**For**  
Consumable such as batteries chargers, rechargeable batteries - 2 years from the date of switch on.  
**Read**  
Consumable such as cables, coils, batteries chargers, rechargeable batteries/battery packs and all others - 1 years from the date of switch on.
- 17. Page No. 22, Point No. 16, Warranty, Sub Point No. D:**  
**For**  
Should quote CMC for two years after warranty.  
**Read**  
**Deleted**
- 18. Page No. 23, In Power Source, Point No. 4:**  
**For**  
Powered by rechargeable cells/Rechargeable battery pack. Recharging equipment & Rechargeable Zinc air cells (at least three or more)/battery pack to be supplied by the manufacturer. Commitment for their free replacement, if they malfunction, within warranty period.  
**Read**  
Powered by rechargeable cells/Rechargeable battery pack. Recharging equipment & Rechargeable cells/battery packs (three or more Sets) to be supplied by the manufacturer. Commitment for their free replacement, if they malfunction, within warranty period. Also provide 10 or more button cells for use in emergency.
- 19. Page No. 23 Point 5, In External Components, Sub Point No. B:**  
**For**  
Two sets of all spares including spare cables and coils and commitment towards free replacement of external components not covered by warranty to be provided for a minimum period of 3 years from date of implantation.  
**Read**  
Two sets of all spares including spare cables and commitment towards free replacement of external components not covered by warranty to be provided for a minimum period of 3 years from date of implantation.
- 20. Page No. 24, Point 16, Warranty, Sub Point No. B:**  
**For**  
External Components, including Speech Processor & battery packs, cables & coils: 3 years from the date of switch on.  
**Read**  
External Components, including Speech Processor: 3 years from the date of switch on.
- 21. Page No. 24, Point 16, Warranty, Sub Point No. C:**  
**For**  
Consumable such as batteries chargers, rechargeable batteries - 2 years from the date of switch on.  
**Read**  
Consumable such as cables, coils, batteries chargers, rechargeable batteries/battery packs and all others - 1 years from the date of switch on.
- 22. Page No. 24, Point 16, Warranty, Sub Point No. D:**  
**For**  
Should quote CMC for two years after warranty.  
**Read**

**Deleted**

**23. Page No. 24, Point No. 18:**

**For**

At least one year of speech and rehabilitative therapy, which would include mapping etc.

**Read**

At least one year of support for speech and rehabilitative therapy, which would include mapping support etc.

**24. Page No. 25, Point No. 4, MIDDLE EAR IMPLANTS, Sub Point No. 09:**

**For**

Power Source: Powered by 675 Zinc air cells

**Read**

Power Source: Powered by 675 Zinc air cells or equivalent (desirable - rechargeable battery with long backup)