

Invitation of quotation
for
Supply of Phototherapy - Single Surface
At
All India Institute of Medical Sciences, Jodhpur

Inquiry No.: : Admin/Gen/03-06/2023-AIIMS.JDH

Inquiry Issue Date : 11th August, 2023

Last Date of Submission : 17th August, 2023 at 03:00 PM.



All India Institute of Medical Sciences, Jodhpur

Basni Phase - II, Jodhpur – 342005, Rajasthan

Telefax: 0291- 2740741, email: procurement@aiimsjodhpur.edu.in

www.aiimsjodhpur.edu.in

Invitation of quotation for Supply of Phototherapy - Single Surface at AIIMS Jodhpur

Sealed Quotations are hereby invited by the undersigned on behalf of the Director, AIIMS Jodhpur for Supply of Phototherapy - Single Surface for the Institute as per terms & conditions mentioned below. The filled quotations along with all the required document must reach in the office of the undersigned on or before 17.08.2023 03:00 PM. The Envelope containing the quotation would please be sealed and super scribed as under:-

“QUOTATION FOR SUPPLY OF PHOTOTHERAPY - SINGLE SURFACE AGAINST INQUIRY NO. ADMN/GEN/03-06/2023-AIIMS.JDH” DUE ON 17.08.2023 03:00 PM”

1. Terms & Conditions:

- A) The quotations received after this deadline & unsealed shall not be entertained under any circumstances whatsoever. In case of postal delay this Institute will not be responsible. **The offer Submitted Fax/Email shall not be considered and no correspondence will be entertained in this matter.**
- B) Quotations must be in the enclosed prescribed Performa on the letter head of the firm duly signed by the Proprietor/ Partner/ Director or their authorized representative, In case of signing of quotation by the authorized representative letter of authorization must be attached with the quotation. Quotation must be dropped in “**Quotation Box**” located in Administration Block of AIIMS, Jodhpur.
- C) Rates must be quoted in **Indian rupees** and as per the format specified taxes extra if any must be written separately.
- D) Rates must be quoted FOR basis (including Freight charges, Insurance, installation etc.)
- E) No overwriting or cutting is permitted in the rate. If found, the quotation shall be summarily rejected.
- F) The rates quoted must be valid for 60 days minimum from the date of opening of the quotation and silence of any tendered on this issue shall be treated as agreed with this condition.
- G) Becoming L1 will not be the criteria for awarding of purchase order unless the rates are reasonable & justified.
- H) RTGS/NEFT details need to be furnished by the supplier with the quotation on the letter head of supplier/firm/agency.
- I) The firm/agency may satisfy the following conditions and attach self-attested copy of the same with the quotation:
- Firm shall be registered with the Government of Rajasthan / Central Government.
 - The firm shall have valid GST No.
 - **The firm should not be black listed by any Govt. Agency/Dept.**
- J) Quotations qualified by such vague and indefinite expressions such as “subject to prior

INQUIRY NO. Admin/Gen/03-06/2023-AIIMS.JDH

confirmation”, “subject to immediate acceptance” etc. will be treated as vague offers and rejected accordingly. Any conditional quotation shall be rejected summarily.

- K) **Delivery Period** – within 30 days from Purchase order.
- L) **Liquidated Damage:** - If the supplier fails to deliver the material on or before the stipulated date, then a penalty at the rate of 0.5 % per week of the total order value shall be levied subject to maximum of 10% of the total order value.
- M) **Payment Terms:** Payment will be only after satisfactorily delivery / commissioning of material and after inspection by the AIIMS Jodhpur.
- N) **Disputes:** -In the event of any dispute or disagreement arising between the contractors and any other department of AIIMS Jodhpur with regards to the interpretation of “Terms & Conditions” of this inquiry, the same shall be referred to the Director, AIIMS Jodhpur whose decision will be final and binding upon the contractor.
- O) AIIMS, Jodhpur reserves the right to increase or decrease quantity and / or amount of work. Decision of Quantity of material in the AIIMS, Jodhpur will be final in this regard.
- P) AIIMS, Jodhpur reserves the right to reject any quotation or part or the whole of inviting quotation process without assigning any reason. Decision of the AIIMS, Jodhpur will be final in this regard.

2. Special Terms & Conditions:

- A) **Bidder must quote the product as per specification provided in Annexure 1.**
- B) **Catalog must be attached with quotation for technical evaluation.**
- C) **The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.**

Deputy Director (Admin)

Encl.: Annexure 1 (Specification)
Annexure 2 (Format of price bid)

Annexure 1

S. No.	Particular	Specifications	Qty.
1.	Phototherapy - Single Surface	<ol style="list-style-type: none"> 1. Blue light LED based (strictly) which should last for at least 30,000 hours 2. Spectrum: peak at 440 to 470 nm, no irradiance in UV or IR ranges (certificate by recognized lab to be produced) 3. Spectral Irradiance of at least $30 \mu\text{W} \cdot \text{cm}^{-2} \cdot \text{nm}^{-1}$ at 45 cm distance between bed and light unit 4. Effective surface area should be at least 20 x 40 cm 5. Two levels of irradiance: <ol style="list-style-type: none"> a) Lower: at least 30 uW / cm² / nm b) Higher: at least 50 uW / cm² / nm 6. Light head should be compact to use along with the radiant warmer & should be provided with tilting facility (at least 90 degree on each side) so that the unit is not coming directly under warmer. 7. Light unit should have white LED's for examination purpose 8. Head height adjustable, approx: 1.30 to 1.75 m 9. Integrated timer for monitoring therapy hours (resettable) & lamp usage hours (non resettable) 10. Sturdy mobile stand 11. The base of the unit should be such that it will go beneath any Incubator/radiant warmer/bed 12. Antistatic castors, 2 with breaks 13. Option of mounting on radiant warmer 14. Option of keeping directly on the roof of incubator 15. Cooling Fan to be provided to dissipate the heat created by LED's 16. Coating: Epoxy/powder coated body for scratch and rust prevention and PU (Poly Urethane) coating for plastic 17. Should have visual and audible alarm indicating the excess of internal temperature and failure of the fan 18. Standards, safety and training <ol style="list-style-type: none"> a) Should be US FDA or CE approved product b) Manufacturer should be ISO certified for quality standards c) Equipment Shall CERTIFIED to be meeting Electrical Safety requirements as per IEC 60601-2-50 Medical Electrical Equipment part-2-50 Particular requirements for the safety of Infant Phoootherapy Equipments d) Should have local service facility. The service provider should have the necessary equipments recommended by the manufacturer to carry out preventive maintenance test as per guidelines provided in the service/maintenance manual e) Training and installation at end user site 19. Warranty <ol style="list-style-type: none"> a) Two years from date of satisfactory installation b) Maintenance contract - CMC for 5 years after warranty 20. Power supply - Power input to be 220-240VAC, 50Hz 21. Items covered under warranty/CMC <ol style="list-style-type: none"> a) Warranty and CMC must cover (but not limited to) all parts including all glass, rubber, metallic, plastic, and electrical/electronic parts. Parts that are not covered must be 	05 Nos.

INQUIRY NO. Admin/Gen/03-06/2023-AIIMS.JDH

		<p>explicitly stated up front.</p> <p>b) Prices of consumables and accessories/spare parts should be quoted separately and the prices should be frozen for the period of warranty and CMC.</p> <p>22. Environmental factors</p> <p>a) The unit shall be capable of being stored continuously in ambient temperature of 0-50deg C and relative humidity of 15-90%</p> <p>b) The unit shall be capable of operating continuously in ambient temperature of 10-40 deg C and relative humidity of 15-90%</p> <p>23. Documentation</p> <p>a) User/Technical/Maintenance manuals to be supplied in English</p> <p>b) Certificate of calibration and inspection from factory</p> <p>c) List of important spares and accessories with their part number and costing</p> <p>d) Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist. The job description of the hospital technician and company service engineer should be clearly spelt out</p> <p>24. User manual and technical manual with trouble shooting guidance in English should be provided</p> <p>25. Company should certify that model quoted is latest and not obsolete, and spares will be available for next 5 years after the completion of warranty.</p> <p>26. Onsite physical demonstration/training of the equipment to all the end users with all the requested facilities will be mandatory</p> <p>27. Service back-up must be available in Jodhpur and complaint must be attended within 72 hours. If down time is >5 days, a replacement unit must be provided failing which an appropriate penalty will be adjusted from warranty/CMC.</p>	
--	--	--	--

Note: -

- ❖ **Catalog/ Brochure must be attached for Technical Evaluation.**
- ❖ The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.

[On the letterhead of firm]

ANNEXURE "2"
PRICE BIDFORM

To,
Deputy Director (Admin),
AIIMS, Jodhpur.

Dear Sir,

1. I/We Submitted the quotation for Enquiry No. "QUOTATION FOR SUPPLY OF PHOTOTHERAPY - SINGLE SURFACE AT AIIMS AGAINST THE INQUIRY NO. Admn/Gen/03-06/2023-AIIMS.JDH" DUE ON 17.08.2023 03:00 PM for Supply of Phototherapy - Single Surface at AIIMS Jodhpur".
2. I/We thoroughly examined, understood and accepted terms & conditions given in the enquiry document, failing which my quotation will be rejected out rightly.
3. I/We hereby offer to supply at the following rates.

S. No	Particular	Qty.	Quoted Make	Price/Unit Exclusive of GST (INR)	GST/Other Taxes	Price/ Unit Inclusive of GST (INR)	Total Cost Inclusive of GST (INR)	MRP
1.	Phototherapy - Single Surface Specification:- As per annexure - 1	05 Nos.						

Note:-

1. **The Bidder must quote only single Make & Model.**
2. The bidder must quoted their quotation only in above said format on the letter of firm otherwise quotation will be **REJECTED**.
3. Catalog must be attached with quotation for technical evaluation.
4. The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.

Date _____

(Name) _____

Place _____

Name of Firm/Company/Agency _____

GSTIN No.: _____

Bank Name:- _____

Bank Account No.: _____

IFSC Code:- _____

Branch Name: _____

Phone No. _____

Email: _____

(Signature of Authorized Person) _____

Seal: _____