

Invitation of quotation
for
Supply of Hospital Grade Electrical Breast Pumps
for multiple user with Accessories tubing's

At

All India Institute of Medical Sciences, Jodhpur

Inquiry No.: : Admin/Gen/72-08/2022-AIIMS.JDH

Inquiry Issue Date : 15th July, 2022

Last Date of Submission : 21st July, 2022 at 03:00 PM.



All India Institute of Medical Sciences, Jodhpur

Basni Phase - II, Jodhpur – 342005, Rajasthan

Telefax: 0291- 2740741, email: procurement@aiimsjodhpur.edu.in

www.aiimsjodhpur.edu.in

Invitation of quotation for Supply of Hospital Grade Electrical Breast Pumps for multiple user with Accessories tubing's at AIIMS Jodhpur

Sealed Quotations are hereby invited by the undersigned on behalf of the Director, AIIMS Jodhpur for Supply of Hospital Grade Electrical Breast Pumps for multiple user with Accessories tubing's for the Institute as per terms & conditions mentioned below. The filled quotations along with all the required document must reach in the office of the undersigned on or before 21.07.2022 03:00 PM. The Envelope containing the quotation would please be sealed and super scribed as under:-

“QUOTATION FOR SUPPLY OF HOSPITAL GRADE ELECTRICAL BREAST PUMPS FOR MULTIPLE USER WITH ACCESSORIES TUBING'S AGAINST INQUIRY NO. ADMN/GEN/72-08/2022-AIIMS.JDH” DUE ON 21.07.2022 03:00 PM

1. Terms & Conditions:

- A) The quotations received after this deadline & unsealed shall not be entertained under any circumstances whatsoever. In case of postal delay this Institute will not be responsible. **The offer Submitted Fax/Email shall not be considered and no correspondence will be entertained in this matter.**
- B) Quotations must be in the enclosed prescribed Performa on the letter head of the firm duly signed by the Proprietor/ Partner/ Director or their authorized representative, In case of signing of quotation by the authorized representative letter of authorization must be attached with the quotation. Quotation must be dropped in “**Quotation Box**” located in Administration Block of AIIMS, Jodhpur.
- C) Rates must be quoted in **Indian rupees** and as per the format specified taxes extra if any must be written separately.
- D) Rates must be quoted FOR basis (including Freight charges, Insurance, installation etc.)
- E) No overwriting or cutting is permitted in the rate. If found, the quotation shall be summarily rejected.
- F) The rates quoted must be valid for 60 days minimum from the date of opening of the quotation and silence of any tendered on this issue shall be treated as agreed with this condition.
- G) Becoming L1 will not be the criteria for awarding of purchase order unless the rates are reasonable & justified.
- H) RTGS/NEFT details need to be furnished by the supplier with the quotation on the letter head of supplier/firm/agency.
- I) The firm/agency may satisfy the following conditions and attach self-attested copy of the same with the quotation:
 - Firm shall be registered with the Government of Rajasthan / Central Government.
 - The firm shall have valid GST No.

INQUIRY NO. Admin/Gen/72-07/2022-AIIMS.JDH

- **The firm should not be black listed by any Govt. Agency/Dept.**

- J) Quotations qualified by such vague and indefinite expressions such as “subject to prior confirmation”, “subject to immediate acceptance” etc. will be treated as vague offers and rejected accordingly. Any conditional quotation shall be rejected summarily.
- K) **Delivery Period** – within 30 days from Purchase order.
- L) **Liquidated Damage:** - If the supplier fails to deliver the material on or before the stipulated date, then a penalty at the rate of 0.5 % per week of the total order value shall be levied subject to maximum of 10% of the total order value.
- M) **Payment Terms:** Payment will be only after satisfactorily delivery / commissioning of material and after inspection by the AIIMS Jodhpur.
- N) **Disputes:** -In the event of any dispute or disagreement arising between the contractors and any other department of AIIMS Jodhpur with regards to the interpretation of “Terms & Conditions” of this inquiry, the same shall be referred to the Director, AIIMS Jodhpur whose decision will be final and binding upon the contractor.
- O) AIIMS, Jodhpur reserves the right to increase or decrease quantity and / or amount of work. Decision of Quantity of material in the AIIMS, Jodhpur will be final in this regard.
- P) AIIMS, Jodhpur reserves the right to reject any quotation or part or the whole of inviting quotation process without assigning any reason. Decision of the AIIMS, Jodhpur will be final in this regard.

2. Special Terms & Conditions:

- A) **Bidder must quote the product as per specification provided in Annexure 1.**
- B) **Catalog must be attached with quotation for technical evaluation.**
- C) **The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.**

Deputy Director (Admin)

Encl.: Annexure 1 (Specification)
Annexure 2 (Format of price bid)

Annexure 1

S. No.	Particular	Specifications	Qty.
1.	Hospital Grade Electrical Breast Pumps for multiple user with Accessories tubing's	<p>SPECIFICATIONS</p> <ol style="list-style-type: none"> 1. It should be a bench top type breast pump. 2. It should be a hospital grade breast pump; capable of working 24X7 in a hospital and can be used by several mothers; one user at a time. 3. It should be easy to operate and provides safe, hassle-free, comfortable and painless pumping experience. 4. It should have a compact portable construction; easily movable. 5. The operation of pump should be automatic. 6. The vacuum needs to be variable and there should be a system by which the vacuum can be adjusted as per the comfort level. 7. The pumping speed needs to be variable. The number of suction cycles per minute needs to be regulated as per requirement. 8. The parts of the pump which comes in direct contact with breast milk should be made of food grade hygienic material which is safe for newborn baby and easy to clean and sterilise. 9. During expression milk will be stored in a bottle attached to the breast shield and there should be no chance of entering milk into the pump-tubing as well as the pump. 10. During expression, the expressed breast milk should not be exposed to the environment. The system must be designed in such a manner so that it will protect the outside air from reaching the expressed milk. 11. During expression, milk will be stored in a bottle and there should be no chance of contamination from the outside environment. 12. The collection system should provide a barrier in between the pump-tubing and the breast shield, which prevents outside air from reaching the expressed milk. 13. Power supply: It should be capable of functioning in power supply of 230V AC, 50Hz. 14. It should be a safety certified approved product (USFDA and CE). 15. Manufactures/ suppliers should have ISO certificate to Quality Standard. 16. Electrically safety conforms to the standards for electrical safety. 17. Accessories/fittings (like tubing, breast shield, connector, valve, membrane, membrane cap, reusable bottles etc.) must be reusable. 18. The equipment should be supplied with 10 (ten) sets of complete accessories/fittings sets (like tubing, breast shield, connector, valve, membrane, membrane cap, reusable bottles etc. which are required for pumping). 19. The equipment should be supplied with all accessories which are required for breast pumping operations. 20. The equipment should be able to function in hot and humid tropical weather condition. 21. Installation: Free delivery & installation and onsite demonstration & training are to be provided. 22. DOCUMENTATION 23. • User/ Technical/ Maintenance manuals to be supplied - should be in English. 24. List of important spare parts and accessories with their part number and 	01 Nos.

INQUIRY NO. Admin/Gen/72-07/2022-AIIMS.JDH

		costing are to be furnished. 25. Costing of all accessories/ fittings those are required for expressing milk is to be furnished. 26. SERVICING & WARRANTY 27. Free servicing of the equipment as per its requirement up to 5 years from the date of installation. 28. Warranty: Minimum 5 years of warranty of all parts of the equipment from the date of installation	
--	--	--	--

Note: -

- ❖ **Catalog/ Brochure must be attached for Technical Evaluation.**
- ❖ The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.

[On the letterhead of firm]

ANNEXURE "2"
PRICE BIDFORM

To,
Deputy Director (Admin),
AIIMS, Jodhpur.

Dear Sir,

1. I/We Submitted the quotation for Enquiry No. "QUOTATION FOR SUPPLY OF HOSPITAL GRADE ELECTRICAL BREAST PUMPS FOR MULTIPLE USER WITH ACCESSORIES TUBING'S AT AIIMS AGAINST THE INQUIRY NO. Admn/Gen/72-08/2022-AIIMS.JDH" DUE ON 21.07.2022 03:00 PM for Supply of Hospital Grade Electrical Breast Pumps for multiple user with Accessories tubing's at AIIMS Jodhpur".
2. I/We thoroughly examined, understood and accepted terms & conditions given in the enquiry document, failing which my quotation will be rejected out rightly.
3. I/We hereby offer to supply at the following rates.

S. No	Particular	Qty.	Quoted Make	Price/Unit Exclusive of GST (INR)	GST/Other Taxes	Price/ Unit Inclusive of GST (INR)	Total Cost Inclusive of GST (INR)	MRP
1.	Hospital Grade Electrical Breast Pumps for multiple user with Accessories tubing's Specification:- As per annexure – 1	01 Pcs.						

Note:-

1. The Bidder must quote only single Make & Model.
2. The bidder must quoted their quotation only in above said format on the letter of firm otherwise quotation will be REJECTED.
3. Catalog must be attached with quotation for technical evaluation.
4. The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.

Date _____

(Name) _____

Place _____

Name of Firm/Company/Agency _____

GSTIN No.: _____

Bank Name:- _____

Bank Account No.: _____

IFSC Code:- _____

Branch Name: _____

Phone No. _____

Email: _____

(Signature of Authorized Person) _____

INQUIRY NO. Admin/Gen/72-07/2022-AIIMS.JDH

Seal: _____