

Invitation of quotation
for
Intramedullary Reaming System-Tibia Surgeries
At
All India Institute of Medical Sciences, Jodhpur

Inquiry No.: : Admin/Gen/79-10/2020-AIIMS.JDH

Inquiry Issue Date : 09th March, 2021

Last Date of Submission : 16th March, 2021 at 03:00 PM.



All India Institute of Medical Sciences, Jodhpur

Basni Phase - II, Jodhpur – 342005, Rajasthan

Telefax: 0291- 2740741, email: procurement@aiimsjodhpur.edu.in

www.aiimsjodhpur.edu.in

Invitation of quotation for Intramedullary Reaming System-Tibia Surgeries at AIIMS Jodhpur

Sealed Quotations are hereby invited by the undersigned on behalf of the Director, AIIMS Jodhpur for Intramedullary Reaming System-Tibia Surgeries for the Institute as per terms & conditions mentioned below. The filled quotations along with all the required document must reach in the office of the undersigned on or before 16.03.2021 03.00 PM. The Envelope containing the quotation would please be sealed and super scribed as under:-

“QUOTATION FOR INTRAMEDULLARY REAMING SYSTEM-TIBIA SURGERIES AGAINST INQUIRY NO. ADMN/GEN/79-10/2020-AIIMS.JDH” DUE ON 16.03.2021 03.00 PM”

1. Terms & Conditions:

- A) The quotations received after this deadline & unsealed shall not be entertained under any circumstances whatsoever. In case of postal delay this Institute will not be responsible. **The offer Submitted Fax/Email shall not be considered and no correspondence will be entertained in this matter.**
- B) Quotations must be in the enclosed prescribed Performa on the letter head of the firm duly signed by the Proprietor/ Partner/ Director or their authorized representative, In case of signing of quotation by the authorized representative letter of authorization must be attached with the quotation. Quotation must be dropped in “**Quotation Box**” located in Administration Block of AIIMS, Jodhpur.
- C) Rates must be quoted in **Indian rupees** and as per the format specified taxes extra if any must be written separately.
- D) Rates must be quoted FOR basis (including Freight charges, Insurance, installation etc.)
- E) No overwriting or cutting is permitted in the rate. If found, the quotation shall be summarily rejected.
- F) The rates quoted must be valid for 60 days minimum from the date of opening of the quotation and silence of any tendered on this issue shall be treated as agreed with this condition.
- G) Becoming L1 will not be the criteria for awarding of purchase order unless the rates are reasonable & justified.
- H) RTGS/NEFT details need to be furnished by the supplier with the quotation on the letter head of supplier/firm/agency.
- I) The firm/agency may satisfy the following conditions and attach self-attested copy of the same with the quotation:
- Firm shall be registered with the Government of Rajasthan / Central Government.
 - The firm shall have valid GST/Other taxes and IT PAN.
 - **The firm should not be black listed by any Govt. Agency/Dept.**
- J) Quotations qualified by such vague and indefinite expressions such as “subject to prior confirmation”, “subject to immediate acceptance” etc. will be treated as vague offers and

INQUIRY NO. Admin/Gen/79-10/2020-AIIMS.JDH

rejected accordingly. Any conditional quotation shall be rejected summarily.

- K) **Delivery Period** – within 30 days from Purchase order.
- L) **Liquidated Damage:** - If the supplier fails to deliver the material on or before the stipulated date, then a penalty at the rate of 0.5 % per week of the total order value shall be levied subject to maximum of 10% of the total order value.
- M) **Payment Terms:** Payment will be only after satisfactorily delivery / commissioning of material and after inspection by the AIIMS Jodhpur.
- N) **Disputes:** -In the event of any dispute or disagreement arising between the contractors and any other department of AIIMS Jodhpur with regards to the interpretation of “Terms & Conditions” of this inquiry, the same shall be referred to the Director, AIIMS Jodhpur whose decision will be final and binding upon the contractor.
- O) AIIMS, Jodhpur reserves the right to increase or decrease quantity and / or amount of work. Decision of Quantity of material in the AIIMS, Jodhpur will be final in this regard.
- P) AIIMS, Jodhpur reserves the right to reject any quotation or part or the whole of inviting quotation process without assigning any reason. Decision of the AIIMS, Jodhpur will be final in this regard.

2. Special Terms & Conditions:

- A) **Bidder must quote the product as per specification provided in Annexure 1.**
- B) **Catalog must be attached with quotation for technical evaluation.**
- C) **The supplier may be asked to arranging demonstration of their Items for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.**

Deputy Director (Admin)

Encl.: Annexure 1 (Specification)
Annexure 2 (Format of price bid)

Annexure 1

S.No.	Particulars	Qty.
1.	7mm diameter flexible shaft should be available which gives a reaming depth to 400-500mm.	1 Each
2.	Reamer Heads starting from 8.5 to 13.5 in increments of 0.5 should be available with 8mm reamer being front cutting.	1 Each
3.	Aluminum Box.	1 Each

Intramedullary Reaming System-Femur	
1.	Company should be at least in its 5 years of operations at the date of submission of tenders in Indian Market.
2.	Bidder must enclose original literatures & technical data sheet in the support of the technical bid.
3.	Physical demonstration should be as per the requirement of user department.
4.	Instruments quality should meet the international standard.
5.	Company should provide material certificates.
6.	<i>Vendor should quote for all the items of same manufacturer</i>
7.	Should be compatible with existing system available in the Institute. (Make:- Depuy Synthes TRS)

Specifications
<ul style="list-style-type: none"> The reaming system intended for reaming large bones and small bones should be able to provide secure reduction.
<ul style="list-style-type: none"> The system should facilitate reduced pressure and minimal heat builds up during its application in surgery.
<ul style="list-style-type: none"> The system should have easy handling and should provide optimal cleaning.
<ul style="list-style-type: none"> The system should comprise of suitable size reamer heads with the NITINOL (super- elastic) shaft.
<ul style="list-style-type: none"> The system should be capable of reaming both clockwise and anticlockwise.
<ul style="list-style-type: none"> The system should come along with reamer diameters of 8.5mm to 13.5mm with .5mm increments.
<ul style="list-style-type: none"> The starting reamer of 8.5mm must be equipped with front cutting edges.
<ul style="list-style-type: none"> The reaming rod should not be less than 900mm and not more than 1200mm with diameter not more than 2.5mm.
<ul style="list-style-type: none"> The shaft should be compatible with the feature of non-manual insertion of the reamer head directly from the tray.
<ul style="list-style-type: none"> The system should come along with the reduction head for fragment reduction with curved and straight tip the displacement of which should not be more than 2.5mm.
<ul style="list-style-type: none"> The reaming system should be made available with the appropriate T-Handle, which should be compatible to hold the shaft manually without the power drive.
<ul style="list-style-type: none"> The system should come along with appropriate size of guide wire for reaming of the large bones and should carry a tissue protector.

Note: - The supplier may be asked to arranging demonstration of their Items for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.

[On the letterhead of firm]

ANNEXURE "2"
PRICE BIDFORM

To,
Administrative Officer,
AIIMS, Jodhpur.

Dear Sir,

1. I/We Submitted the quotation for Enquiry No. "QUOTATION FOR INTRAMEDULLARY REAMING SYSTEM-TIBIA SURGERIES AT AIIMS AGAINST THE INQUIRY NO. Admin/Gen/79-10/2020-AIIMS.JDH" DUE ON 16.03.2021 03.00 PM for Intramedullary Reaming System-Tibia Surgeries at AIIMS Jodhpur".
2. I/We thoroughly examined, understood and accepted terms & conditions given in the enquiry document, failing which my quotation will be rejected out rightly.
3. I/We hereby offer to supply at the following rates.

S. No	Particular	Qty.	Quoted Make	Price/Unit Exclusive of GST (INR)	GST/ Other Taxes	Price/ Unit Inclusive of GST (INR)	Total Cost Inclusive of GST (INR)	MRP
1.	7mm diameter flexible shaft should be available which gives a reaming depth to 400-500mm.	1 Each						
2.	Reamer Heads starting from 8.5 to 13.5 in increments of 0.5 should be available with 8mm reamer being front cutting.	1 Each						
3.	Aluminum Box.	1 Each						

INQUIRY NO. Admin/Gen/79-10/2020-AIIMS.JDH

Note:-

- 1. The Bidder must quote single Make.**
2. The bidder must quote their quotation only in above said format on the letter of firm otherwise quotation will be **REJECTED**.
3. Catalog must be attached with quotation for technical evaluation.
4. The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.

Date _____

Place _____

(Name) _____

Name of Firm/Company/Agency _____

GSTIN No.: _____

Bank Name:- _____

Bank Account No.: _____

IFSC Code:- _____

Branch Name: _____

Phone No. _____

Email: _____

(Signature of Authorized Person) _____

Seal: _____