

Invitation of quotation

for

Multipara Monitor

At

All India Institute of Medical Sciences, Jodhpur

Inquiry No.: : Admin/Gen/11-02/2020-AIIMS.JDH

Inquiry Issue Date : 18th December, 2020

Last Date of Submission : 24th December, 2020 at 03:00 PM.



All India Institute of Medical Sciences, Jodhpur

Basni Phase - II, Jodhpur – 342005, Rajasthan

Telefax: 0291- 2740741, email: procurement@aiimsjodhpur.edu.in

www.aiimsjodhpur.edu.in

Invitation of quotation for Multipara Monitor at AIIMS Jodhpur

Sealed Quotations are hereby invited by the undersigned on behalf of the Director, AIIMS Jodhpur for Multipara Monitor for the Institute as per terms & conditions mentioned below. The filled quotations along with all the required document must reach in the office of the undersigned on or before 24.12.2020 03.00 PM. The Envelope containing the quotation would please be sealed and super scribed as under:-

“QUOTATION FOR MULTIPARA MONITOR AGAINST INQUIRY NO. ADMN/GEN/11-02/2020-AIIMS.JDH” DUE ON 24.12.2020 03.00 PM”

1. Terms & Conditions:

- A) The quotations received after this deadline & unsealed shall not be entertained under any circumstances whatsoever. In case of postal delay this Institute will not be responsible. **The offer Submitted Fax/Email shall not be considered and no correspondence will be entertained in this matter.**
- B) Quotations must be in the enclosed prescribed Performa on the letter head of the firm duly signed by the Proprietor/ Partner/ Director or their authorized representative, In case of signing of quotation by the authorized representative letter of authorization must be attached with the quotation. Quotation must be dropped in “**Quotation Box**” located in Administration Block of AIIMS, Jodhpur.
- C) Rates must be quoted in **Indian rupees** and as per the format specified taxes extra if any must be written separately.
- D) Rates must be quoted FOR basis (including Freight charges, Insurance, installation etc.)
- E) No overwriting or cutting is permitted in the rate. If found, the quotation shall be summarily rejected.
- F) The rates quoted must be valid for 60 days minimum from the date of opening of the quotation and silence of any tendered on this issue shall be treated as agreed with this condition.
- G) Becoming L1 will not be the criteria for awarding of purchase order unless the rates are reasonable & justified.
- H) RTGS/NEFT details need to be furnished by the supplier with the quotation on the letter head of supplier/firm/agency.
- I) The firm/agency may satisfy the following conditions and attach self-attested copy of the same with the quotation:
- Firm shall be registered with the Government of Rajasthan / Central Government.
 - The firm shall have valid GST/Other taxes and IT PAN.
 - **The firm should not be black listed by any Govt. Agency/Dept.**
- J) Quotations qualified by such vague and indefinite expressions such as “subject to prior confirmation”, “subject to immediate acceptance” etc. will be treated as vague offers and rejected accordingly. Any conditional quotation shall be rejected summarily.

INQUIRY NO. Admin/Gen/11-02/2020-AIIMS.JDH

- K) **Delivery Period** – within 30 days from Purchase order.
- L) **Liquidated Damage:** - If the supplier fails to deliver the material on or before the stipulated date, then a penalty at the rate of 0.5 % per week of the total order value shall be levied subject to maximum of 10% of the total order value.
- M) **Payment Terms:** Payment will be only after satisfactorily delivery / commissioning of material and after inspection by the AIIMS Jodhpur.
- N) **Disputes:** -In the event of any dispute or disagreement arising between the contractors and any other department of AIIMS Jodhpur with regards to the interpretation of “Terms & Conditions” of this inquiry, the same shall be referred to the Director, AIIMS Jodhpur whose decision will be final and binding upon the contractor.
- O) AIIMS, Jodhpur reserves the right to increase or decrease quantity and / or amount of work. Decision of Quantity of material in the AIIMS, Jodhpur will be final in this regard.
- P) AIIMS, Jodhpur reserves the right to reject any quotation or part or the whole of inviting quotation process without assigning any reason. Decision of the AIIMS, Jodhpur will be final in this regard.

2. **Special Terms & Conditions:**

- A) **Bidder must quote the product as per specification provided in Annexure 1.**
- B) **Catalog must be attached with quotation for technical evaluation.**
- C) **The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.**

Administrative Officer

Encl.: Annexure 1 (Specification)
Annexure 2 (Format of price bid)

Annexure 1

Sr.No.	Item Name	Qty.	Make
1.	Multipara Monitor	04 Nos.	BPL/Skanray/Drager

Specifications of Multipara Monitor:-

Suitable for Usage in Operation Room and ICU Capable of monitoring ECG, SPO2, Non Invasive Blood Pressure(NIBP), Respiration Rate and Temperature up gradable to IBP & ETCO2

- Should have large colour display of 10.4 inch screen
- Device should be compact, portable and light-weight
- Should display 10 waveforms or more
- Waveforms channels should be user selectable
- Should be capable of displaying waveforms and numeric values simultaneously
- User Selectable colour for waveforms
- Should have various selectable display modes like Big Font, Oxy Crg, Short Trend, NiBp Review, 7-Lead ECG
- Should be able to monitor ECG, SpO2, NiBp, Respiration and Temperature
- **ECG:**
- Should be able to monitor ECG through 5-Lead Patient Cable
- Should be able to display Lead I,II,III, aVR, aVL, aVF and one of the chest leads
- Should be able to monitor Heart Rate from 15-350 bpm
- Should have an interface for displaying all the ECG Leads Monitored
- Should have user selectable modes like Monitor, Surgery and Diagnostic for operation in ICU, OT's etc
- Should have pacemaker detection facility
- Should have arrhythmia detection
- **SPO2:**
- Should use Digital technology for monitor SpO2
- Should display the numeric value and plethysmograph as well
- Should display the value from 1-100%
- **TEMPERATURE**
- Should be able to monitor dual temperature values
- Should display the rectal as well as skin temperature
- Should also display the difference between these values
- **NiBp:**
- Should follow the automatic oscillometric method for measurement of Nibp
- Should have cuffs for adult, paediatric and neonatal patients
- Should have a measuring range of 0-300 mmHg
- Should have auto, manual and stat modes of operation
- **RESPIRATION**
- Should display numeric values and respiration waveform as well
- Should have the option for selection of lead for monitoring Respiration
- Should have apnea detection facility
- Should have upto 120 seconds for apnea detection

INQUIRY NO. Admin/Gen/11-02/2020-AIIMS.JDH

- Should have alarm settings for all parameters
- Should have color coded alarms for different priority of alarms
- Should have audio and visual alarms
- Should have separate volume control and QRS Beep Volume
- Should have short-cut keys for freeze, Nibp, Alarm Mute
- Should have navigator wheel for easy selection of parameters
- Should have drug dose and OxyCrg Interfaces
- Should have Heart Rate Variability interface for providing with beat-beat analysis of patient heart rate
- Should be portable and handle should be provided
- Should have the following data storage interfaces:
 - Should have a maximum of 168 Hours of Trend Data storage
 - Should be able store 1000-1800 alarm events
 - Should have 750-1000 NiBp Measurement values storage option
- Should have lead acid battery with provision for adding a second battery
- Should work on 220V, 50 Hz supply
- Should have CNS connectivity Ready.
- Device Should be CE certified
- Manufacturer should conform to ISO Standards
- Should have local service centre presence / service facility.
- Warranty- One year on main unit.
- The device should be supplied with standard accessories:
 - 5 Lead ECG Cable: 1 No
 - Disposable ECG Electrodes: 1 Bag
 - SpO2 Probe: 1 No
 - NiBp Cuffs: 3 Sets
 - Temperature Probe: 1 No
 - Power Cord: 1 No

Note: -The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.

[On the letterhead of firm]

ANNEXURE "2"
PRICE BIDFORM

To,
Administrative Officer,
AIIMS, Jodhpur.

Dear Sir,

1. I/We Submitted the quotation for Enquiry No. "QUOTATION FOR MULTIPARA MONITOR AT AIIMS AGAINST THE INQUIRY NO. Admn/Gen/11-02/2020-AIIMS.JDH" DUE ON 24.12.2020 03.00 PM for Multipara Monitor at AIIMS Jodhpur".
2. I/We thoroughly examined, understood and accepted terms & conditions given in the enquiry document, failing which my quotation will be rejected out rightly.
3. I/We hereby offer to supply at the following rates.

S. No	Particular	Qty.	Quoted Make	Price/Unit Exclusive of GST (INR)	GST/ Other Taxes	Price/ Unit Inclusive of GST (INR)	Total Cost Inclusive of GST (INR)	MRP
1.	Multipara Monitor	04 Nos.						

Note:-

1. **The Bidder must quote single Make.**
2. The bidder must quoted their quotation only in above said format on the letter of firm otherwise quotation will be **REJECTED**.
3. Catalog must be attached with quotation for technical evaluation.
4. The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.

Date _____

(Name) _____

Place _____

Name of Firm/Company/Agency _____

GSTIN No.: _____

Bank Name:- _____

Bank Account No.: _____

IFSC Code:- _____

Branch Name: _____

Phone No. _____

Email: _____

(Signature of Authorized Person) _____

Seal: _____