



All India Institute of Medical Sciences,
Jodhpur

Dated: 24th August, 2020

CORRIGENDUM

Invitation of quotation

for

Portable Suction Machine

At

All India Institute of Medical Sciences, Jodhpur

Inquiry No.: : Admin/Gen/60-01/2020-AIIMS.JDH

Inquiry Issue Date : 22nd August, 2020

Last Date of Submission : 28th August, 2020 at 05:00 PM.

The following revisions / modifications is made –

Subject- Changes in Qty. (Portable Suction Machine-01 Nos.)

Sr No.	Item Name	Specification	Qty.
1.	Portable Suction Machine	<ul style="list-style-type: none">• Plastic Moulded on two wheels with foot switch & change over Lever• Max Vacuum -700mm Hg , LPM:90L/Min• Double Piston Pump• Jars : 2 x 2.5 L PC Jar• Bacterial Filter Autoclavable/Reusable• Tubing : 10mm ID X 2 Mtr (PVC)• Vacuum Guage : 6.25 cm Dia 0-760mmHg• Power Supply: 220V AC,50/60Hz,580Watt• Noise Level : 55 Db A +/-3• Overflow Safety :Mechanical & Electronic Sensor.	01 Nos.