



**ALL INDIA INSTITUTE OF MEDICAL SCIENCES,
JODHPUR**

Inquiry No.: Admn/General/43/2015-AIIMS.JDH

Date: - 17.07.2015

Invitation of quotation for Dental Scaler for AIIMS Jodhpur

Sealed Quotations are hereby invited by the undersigned on behalf of the Director, AIIMS Jodhpur for supply of Dental Scaler for the Institute as per terms & conditions mentioned below. The filled quotations along with all the required document must reach in the office of the undersigned on or before 24.07.2015 05.00 PM. The Envelope containing the quotation would please be sealed and super scribed as under:-

**“QUOTATION FOR DENTAL SCALER AGAINST INQUIRY NO.
Admn/General/43/2015-AIIMS.JDH” DUE ON 24.07.2015 05.00 PM”**

1. Terms & Conditions:

- A) The quotations received after this deadline & unsealed shall not be entertained under any circumstances whatsoever. In case of postal delay this Institute will not be responsible.
- B) Quotations must be in the enclosed prescribed Performa on the letter head of the firm duly signed by the Proprietor/ Partner/ Director or their authorized representative, In case of signing of quotation by the authorized representative letter of authorization must be attached with the quotation.
- C) Rates must be quoted as per the format specified Taxes extra if any must be written separately.
- D) Rates must be quoted FOR basis (including Freight charges, Insurance etc.)
- E) No overwriting or cutting is permitted in the rate. If found, the quotation shall be summarily rejected.
- F) The rates quoted must be valid for 60 days minimum from the date of opening of the quotation and silence of any tendered on this issue shall be treated as agreed with this condition.
- G) Becoming L1 will not be the criteria for awarding of purchase order unless the rates are reasonable & justified.
L1 will be decided on individual item basis.
- H) RTGS/NEFT details need to be furnished by the supplier with the quotation on the letter head of supplier/firm/agency.
- I) The firm/agency may satisfy the following conditions and attach self-attested copy of the same with the quotation:
 - Firm shall be registered with the Government of Rajasthan / Central Government.
 - The firm shall have valid VAT/ Sales Tax No. and IT PAN.
 - **The firm should not be black listed by any Govt. Agency/Dept.**



ALL INDIA INSTITUTE OF MEDICAL SCIENCES, JODHPUR

- J) Quotations qualified by such vague and indefinite expressions such as “subject to prior confirmation”, “subject to immediate acceptance” etc. will be treated as vague offers and rejected accordingly. Any conditional quotation shall be rejected summarily.
- K) **The supplier may be asked to given a demonstration of the product(s), which will be evaluated by a technical evaluation committee. The expenditure incurred for demonstrating the items will be borne by the supplier.**
- L) **Delivery Period** – 15 days from award of work.
- M) **Liquidated Damage:** - If the supplier fails to deliver the material on or before the stipulated date, then a penalty at the rate of 0.5 % per week of the total order value shall be levied subject to maximum of 10% of the total order value.
- N) **Payment Terms:** Payment will be only after satisfactorily delivery / commissioning of material and after inspection by the AIIMS Jodhpur.
- O) **Disputes:** -In the event of any dispute or disagreement arising between the contractors and any other department of AIIMS Jodhpur with regards to the interpretation of “Terms & Conditions” of this inquiry, the same shall be referred to the Director, AIIMS Jodhpur whose decision will be final and binding upon the contractor.
- P) AIIMS, Jodhpur reserves the right to increase or decrease quantity and / or amount of work. Decision of Quantity of material in the AIIMS, Jodhpur will be final in this regard.
- Q) AIIMS, Jodhpur reserves the right to reject any quotation or part or the whole of inviting quotation process without assigning any reason. Decision of the AIIMS, Jodhpur will be final in this regard.

Administrative Officer

Encl.:

1. Annexure 1 (Format of Price Bid)
2. Specification for Dental Scaler



**ALL INDIA INSTITUTE OF MEDICAL SCIENCES,
JODHPUR**

[On the letterhead of firm]

ANNEXURE "2"

PRICE BIDFORM

To,

Administrative Officer,
AIIMS Jodhpur.

Dear Sir,

1. I/We Submitted the quotation for Enquiry No. **"QUOTATION FOR DENTAL SCALER AGAINST THE INQUIRY NO. Admn/General/43/2015-AIIMS.JDH"** DUE ON 24.07.2015 05.00 PM for Supply of Dental Scaler at AIIMS Jodhpur".
2. I/We thoroughly examined, understood and accepted terms & conditions given in the enquiry document, failing which my quotation will be rejected out rightly.
3. I/We hereby offer to supply at the following rates.

S. No.	Particulars	Quantity	Make	Price/Unit Exclusive of TAX (INR)	TAX %	Price/Unit Inclusive of TAX	Total Amount Inclusive of TAX	MRP
1	Dental Scaler	1						

Date _____

Place _____

(Signature of Authorized Person) _____

(Name) _____

Name of Firm/Company/Agency _____

Phone No. _____

Email: _____



ALL INDIA INSTITUTE OF MEDICAL SCIENCES, JODHPUR

SPECIFICATIONS FOR DENTAL SCALER

1 Description of Function

1.1 Dental Scaler is required for removing the supragingival and subgingival calculus.

2 Operational Requirements

2.1 Microprocessor based system

3 Technical Specifications

3.1 Based on piezoelectric technology

3.2 Having torque tool for tightening of the tip

3.3 High power turbo mode and low power mode

3.4 Should have LED light in scaler handpiece

3.5 Automatic smart power feedback control

3.6 Minimum vibration frequency of 28-36 KHZ and

3.7 Twenty tips for scaler, one endodontic ultrasonic kit and one set of perio-curette tips

3.8 Foot pedal

3.9 Separate control for water and tip vibration

3.10 Should be supplied with two autoclavable hand pieces.

3.11 It should have self-contained tank of 300 ml capacity

4 System Configuration Accessories, spares and consumables

4.1 System as specified

5 Environmental factors

5.1 The unit shall be capable of being stored continuously in ambient temperature of 0 -50deg C and relative humidity of 15-90%

5.2 The unit shall be capable of operating in ambient temperature of 20-30 deg C and relative humidity of less than 70%

6 Power Supply

6.1 Power input to be 220-240VAC, 50Hz fitted with Indian plug

7 Standards, Safety and Training

7.1 Should comply with Medical Device class II type BF, in conformity to the requisites of Directive 93/842/CEE for the SCALER unit

7.2 Should be FDA/ CE approved product

7.3 Electrical safety conforms to standards for electrical safety IEC-60601 / IS-13450

8 Documentation

8.1 User/Technical/Maintenance manuals to be supplied in English.

8.2 Certificate of calibration and inspection.

8.3 List of important spare parts and accessories with their part number and costing