

**Invitation of quotation**  
**for**  
**Electrical Items for STP Pump Room**  
**At**  
**All India Institute of Medical Sciences, Jodhpur**

Inquiry No.: : Admn/General/25/2015-AIIMS.JDH

Inquiry Issue Date : 29<sup>th</sup> December, 2015

Last Date of Submission : 04<sup>th</sup> January, 2016 at 05:00 PM.



**All India Institute of Medical Sciences, Jodhpur**

Basni Phase - II, Jodhpur – 342005, Rajasthan

Telefax: 0291- 2012978, email: [procurement@aiimsjodhpur.edu.in](mailto:procurement@aiimsjodhpur.edu.in)

[www.aiimsjodhpur.edu.in](http://www.aiimsjodhpur.edu.in)

**Invitation of quotation for Electrical Items for STP Pump Room at  
AIIMS Jodhpur**

Sealed Quotations are hereby invited by the undersigned on behalf of the Director, AIIMS Jodhpur for supply of Consumable Items for the Institute as per terms & conditions mentioned below. The filled quotations along with all the required document must reach in the office of the undersigned on or before 04.01.2016 05.00 PM. The Envelope containing the quotation would please be sealed and super scribed as under:-

**“QUOTATION FOR Electrical Items for STP Pump Room AGAINST INQUIRY NO.  
Admn/General/25/2015-AIIMS.JDH” DUE ON 04.01.2016 05.00 PM”**

**1. Terms & Conditions:**

- A) The quotations received after this deadline & unsealed shall not be entertained under any circumstances whatsoever. In case of postal delay this Institute will not be responsible.
- B) Quotations must be in the enclosed prescribed Performa on the letter head of the firm duly signed by the Proprietor/ Partner/ Director or their authorized representative, In case of signing of quotation by the authorized representative letter of authorization must be attached with the quotation.
- C) Rates must be quoted as per the format specified Taxes extra if any must be written separately.
- D) Rates must be quoted FOR basis (including Freight charges, Insurance etc.)
- E) No overwriting or cutting is permitted in the rate. If found, the quotation shall be summarily rejected.
- F) The rates quoted must be valid for 60 days minimum from the date of opening of the quotation and silence of any tendered on this issue shall be treated as agreed with this condition.
- G) RTGS/NEFT details need to be furnished by the supplier with the quotation on the letter head of supplier/firm/agency.
- H) The firm/agency may satisfy the following conditions and attach self-attested copy of the same with the quotation:
  - Firm shall be registered with the Government of Rajasthan / Central Government.
  - The firm shall have valid VAT/ Sales Tax No. and IT PAN.
  - **The firm should not be black listed by any Govt. Agency/Dept.**
- I) Quotations qualified by such vague and indefinite expressions such as “subject to prior confirmation”, “subject to immediate acceptance” etc. will be treated as vague offers and rejected accordingly. Any conditional quotation shall be rejected summarily.

INQUIRY NO. Admn/General/25/2015-AIIMS.JDH

- J) Becoming L1 will not be the criteria for awarding of purchase order unless the rates are reasonable & justified.  
**(L1 will be decided on composite basis)**
- A) The supplier may be asked to submit a sample/catalog of the product(s), which will be evaluated by a technical evaluation committee. The expenditure incurred for demonstrating the items will be borne by the supplier.
- B) **Delivery Period** – 15 days from award of work.
- C) **Liquidated Damage:** - If the supplier fails to deliver the material on or before the stipulated date, then a penalty at the rate of 0.5 % per week of the total order value shall be levied subject to maximum of 10% of the total order value.
- D) **Payment Terms:** Payment will be only after satisfactorily delivery / commissioning of material and after inspection by the AIIMS Jodhpur.
- E) **Disputes:** -In the event of any dispute or disagreement arising between the contractors and any other department of AIIMS Jodhpur with regards to the interpretation of “Terms & Conditions” of this inquiry, the same shall be referred to the Director, AIIMS Jodhpur whose decision will be final and binding upon the contractor.
- F) AIIMS, Jodhpur reserves the right to increase or decrease quantity and / or amount of work. Decision of Quantity of material in the AIIMS, Jodhpur will be final in this regard.
- G) AIIMS, Jodhpur reserves the right to reject any quotation or part or the whole of inviting quotation process without assigning any reason. Decision of the AIIMS, Jodhpur will be final in this regard.

**Administrative Officer**

Encl.:

1. Annexure 1 (Format of Price Bid).

**[On the letterhead of firm]****ANNEXURE "1"**  
**PRICE BID FORM**

To,  
Administrative Officer,  
AIIMS Jodhpur.

Dear Sir,

1. I/We ..... Submitted the quotation for Enquiry No. **"QUOTATION FOR ELECTRICAL ITEMS FOR STP PUMP ROOM AGAINST THE INQUIRY NO. Admn/General/25/2015-AIIMS.JDH" DUE ON 04.01.2016 05.00 PM** for Supply of Electrical Items at AIIMS Jodhpur".
2. I/We thoroughly examined, understood and accepted terms & conditions given in the enquiry document, failing which my quotation will be rejected out rightly.
3. I/We hereby offer to supply at the following rates.

S. No.	Product Name	Recommended Make	Qty.	Quoted Make	Rate Exclusive Taxes per unit	Taxes (if any)	Rate per Unit Inclusive Taxes	Total Cost Inclusive Taxes
1	Air break starter panel made out of sheet steel powder coated enclosure comprising of over load protection relay, short circuit & single phasing protection, ON/OFF push buttons, ammeter, voltmeter, indicating lamps etc. complete in all respect suitable for 3-phase, 415 Volt, 3 HP, Motor Etc.	ADVANCE/ L&T / JYOTI / or Equivalent ISI Make	2 Nos.					
2	Recessed/ Surface mounting heavy duty horizontal type sheet steel Distribution board phophatised/ powder painted complete with suitable rating insulated copper bus bar, shorting link, earth link and din bar, masking sheet	LEGRAND/ HAVELLS / SCHNEIDER	1 Nos.					
3	3 - Phase, 415 Volt, 40Amps, 4-Pole Isolator	LEGRAND/ HAVELLS / SCHNEIDER	1 Nos.					
4	3 - Phase, 415 Volt, 16Amps, TP MCB	LEGRAND/ HAVELLS / SCHNEIDER	2 Nos.					
5	1 - Phase, 230 Volt, 16Amps, SP MCB	LEGRAND/ HAVELLS / SCHNEIDER	4 Nos.					
6	4 Core x 2.5 sqmm FRLS PVC insulated & unsheathed flexible copper conductor ISI marked (IS:694) of 1.1 Kv grade	POLYCAB/ RR Cable	50 Mtr.					
<b>Total Amount in figure (Inclusive of all taxes)</b>								
<b>Total Amount in words (Inclusive of all taxes)</b>								

**(L1 will be decided on composite basis)**

Date \_\_\_\_\_

Place \_\_\_\_\_

(Signature of Authorized Person)

(Name) \_\_\_\_\_

Name of Firm/Company/Agency \_\_\_\_\_

Phone No. \_\_\_\_\_

Email: \_\_\_\_\_