

Invitation of quotation

for

Self - Inflating Bags

At

All India Institute of Medical Sciences, Jodhpur

Inquiry No.: : Admn/General/54/2015-AIIMS.JDH

Inquiry Issue Date : 21st December, 2015

Last Date of Submission : 28th December, 2015 2015 at 05:00 PM.



All India Institute of Medical Sciences, Jodhpur

Basni Phase - II, Jodhpur – 342005, Rajasthan

Telefax: 0291- 2012978, email: procurement@aiimsjodhpur.edu.in

www.aiimsjodhpur.edu.in

Invitation of quotation for Self-Inflating Bags at AIIMS Jodhpur

Sealed Quotations are hereby invited by the undersigned on behalf of the Director, AIIMS Jodhpur for supply of Self Inflating Bags for the Institute as per terms & conditions mentioned below. The filled quotations along with all the required document must reach in the office of the undersigned on or before 28.12.2015 05.00 PM. The Envelope containing the quotation would please be sealed and super scribed as under:-

“QUOTATION FOR SELF-INFLATING BAGS AGAINST INQUIRY NO. Admn/General/54/2015-AIIMS.JDH” DUE ON 28.12.2015 05.00 PM”

1. Terms & Conditions:

- A) The quotations received after this deadline & unsealed shall not be entertained under any circumstances whatsoever. In case of postal delay this Institute will not be responsible.
- B) Quotations must be in the enclosed prescribed Performa on the letter head of the firm duly signed by the Proprietor/ Partner/ Director or their authorized representative, In case of signing of quotation by the authorized representative letter of authorization must be attached with the quotation.
- C) Rates must be quoted as per the format specified Taxes extra if any must be written separately.
- D) Rates must be quoted FOR basis (including Freight charges, Insurance etc.)
- E) No overwriting or cutting is permitted in the rate. If found, the quotation shall be summarily rejected.
- F) The rates quoted must be valid for 60 days minimum from the date of opening of the quotation and silence of any tendered on this issue shall be treated as agreed with this condition.
- G) RTGS/NEFT details need to be furnished by the supplier with the quotation on the letter head of supplier/firm/agency.
- H) The firm/agency may satisfy the following conditions and attach self-attested copy of the same with the quotation:
 - Firm shall be registered with the Government of Rajasthan / Central Government.
 - The firm shall have valid VAT/ Sales Tax No. and IT PAN.
 - **The firm should not be black listed by any Govt. Agency/Dept.**
- I) Quotations qualified by such vague and indefinite expressions such as “subject to prior confirmation”, “subject to immediate acceptance” etc. will be treated as vague offers and rejected accordingly. Any conditional quotation shall be rejected summarily.

- J) Becoming L1 will not be the criteria for awarding of purchase order unless the rates are reasonable & justified.
(L1 will be decided on composite basis)
- K) The supplier may be asked to submit a sample/catalog of the product(s), which will be evaluated by a technical evaluation committee. The expenditure incurred for demonstrating the items will be borne by the supplier.
- L) **Delivery Period** – 15 days from award of work.
- M) **Liquidated Damage:** - If the supplier fails to deliver the material on or before the stipulated date, then a penalty at the rate of 0.5 % per week of the total order value shall be levied subject to maximum of 10% of the total order value.
- N) **Payment Terms:** Payment will be only after satisfactorily delivery / commissioning of material and after inspection by the AIIMS Jodhpur.
- O) **Disputes:** -In the event of any dispute or disagreement arising between the contractors and any other department of AIIMS Jodhpur with regards to the interpretation of “Terms & Conditions” of this inquiry, the same shall be referred to the Director, AIIMS Jodhpur whose decision will be final and binding upon the contractor.
- P) AIIMS, Jodhpur reserves the right to increase or decrease quantity and / or amount of work. Decision of Quantity of material in the AIIMS, Jodhpur will be final in this regard.
- Q) AIIMS, Jodhpur reserves the right to reject any quotation or part or the whole of inviting quotation process without assigning any reason. Decision of the AIIMS, Jodhpur will be final in this regard.

Administrative Officer

Encl.:

1. Annexure 1 (Format of Price Bid).
2. Annexure 2 (Technical Specification).

[On the letterhead of firm]

ANNEXURE "1"
PRICE BID FORM

To,
Administrative Officer,
AIIMS Jodhpur.

Dear Sir,

1. I/We Submitted the quotation for Enquiry No. **"QUOTATION FOR SELF INFLATING BAGS AGAINST THE INQUIRY NO. Admn/General/54/2015-AIIMS.JDH" DUE ON 28.12.2015 05.00 PM** for Supply of Self Inflating Bags at AIIMS Jodhpur".
2. I/We thoroughly examined, understood and accepted terms & conditions given in the enquiry document, failing which my quotation will be rejected out rightly.
3. I/We hereby offer to supply at the following rates.

S. No.	Particulars	Capacity	Qty	Make	Price/Unit Exclusive of TAX (INR)	TAX %	Price/Unit Inclusive of TAX	Total Amount Inclusive of TAX	MRP
1	Self- Inflating Bags	250 ML	8						
2		500 ML	8						
3		1500 ML	8						
Total Amount in figure (Inclusive of all taxes)									
Total Amount in words (Inclusive of all taxes)									

Note:-

1. L1 will be decided on composite basis.
2. The Bidder must be submit the sample/catalog of the quoted make & Model.
3. The Product must be European CE OR USFDA Approved.
4. The Bidder must be submit the European CE OR USFDA approved certificates
5. The Bidder must be quote only single make & model for each items.

Date_____

Place_____

(Signature of Authorized Person)

(Name)_____

Name of Firm/Company/Agency_____

Phone No. _____

Email: _____

ANNEXURE "2"

Technical Specification for Self-Inflating Bag

1. Self-inflating Bag.
2. Silicone made.
3. Provided with closed ended reservoir with two valves.
4. Patient valve pliable, well-sealed, have minimum dead space and no forward or backward leaks.
5. The bag should have an oxygen inlet which fits into the standard oxygen tubing both from a cylinder and central supply.
6. Face masks should be transparent, fit the patient outlet easily and have minimum dead space.
7. The system should withstand washing scrubbing and autoclaving procedures.
8. Face masks: 3 Size i.e. 00,01,02,:3 set with each bag.
9. Should be provided with one set of oxygen tubing and reservoir with each self - inflating bag.
10. Should have a safety valve to control maximum PIP.
11. Option to control PEEP (Optional).
12. European CE **OR** USFDA approved and certificate should be provided. A maximum of 15% variation is allowed in the volumes of self – inflating bags of different volumes as mentioned below:-
 - I. 250 ML – 8
 - II. 500 ML – 8
 - III. 500 ML – 8