



अखिल भारतीय आयुर्विज्ञान संस्थान, जोधपुर
ALL INDIA INSTITUTE OF MEDICAL SCIENCES, JODHPUR

Dated: 24th July, 2021

CORRIGENDUM

Invitation of quotation
for
Skin Grafting Handle and Attached Blades.
At

All India Institute of Medical Sciences, Jodhpur

Inquiry No.: : Admin/Gen/35-01(ii)/2021-AIIMS.JDH

Inquiry Issue Date : 20th July, 2021

Last Date of Submission : 28th July, 2021 at 03:00 PM.



All India Institute of Medical Sciences, Jodhpur

Basni Phase - II, Jodhpur – 342005, Rajasthan

Telefax: 0291- 2740741, email: procurement@aiimsjodhpur.edu.in

www.aiimsjodhpur.edu.in

The following revisions/modification is made –

Subject- Changes in Quantity-

Sr. No.	Item Name	Quantity
1.	Skin Grafting Handle (Down/Watson Type) Specification: As per quotation.	02
2.	Skin Grafting Blades Specification: As per quotation. (Pack of 10 Blade)	02

Specifications for skin grafting handle watson type

- Skin graft handle length should be 300 mm, 12.
- Blade and handle should be quoted of same manufacturer
- Instrument should be made of high-grade stainless steel. The quality of steel should comply with the DIN standards.
- It should be of high quality and precision.
- It should be light weight, strong and durable.
- It should be non-magnetic.
- Instrument should be autoclavable.
- The Tungsten Carbide inlay should be nickel based.
- The TC inlay should be welded to the instrument and not pasted.
- Instrument should have bar coding.
- All instruments should be supplied from single manufacturer.
- The manufacturer should have a direct repair facility available in India.
- All instruments should be autoclavable at 134 deg.
- Instruments should meet the International Certification as per DIN EN ISO 13845:2001.
- It should have TUV certificate of manufacturer with details of manufacturing facility.
- The instruments should have bar coding for individual tracking of the instruments.
- Instruments should be (European CE and US-FDA) OR (US-FDA AND BIS) CERTIFIED.
- If quoted by other than OEM (Original equipment manufacturer), Tender specific Authorization Letter of OEM is mandatory.
- L1- Overall Basis