



Date: -16<sup>th</sup> May, 2017

Corrigendum  
For  
Tender for  
Morris water maze automated with software for the  
Department of Pharmacology

NIT Issue Date	: 03 <sup>rd</sup> April, 2017
NIT No.	: Admn/Tender/16/2017-AIIMS.JDH
Pre-Bid Meeting	: 12 <sup>th</sup> April, 2017 at 03:00 PM
Earlier Last Date of Submission	: 12 <sup>th</sup> May, 2017 at 03:00 PM
Extended Last Date of Submission	: 30 <sup>th</sup> May, 2017 at 03:00 PM
Bid opening	: 31 <sup>st</sup> May, 2017 at 03:30 PM

**The following revised and additional specification will be added:-**

**1. Page No. 11, In technical specification, Line No. 23:**

**For**

Connectivity with any video tracking system that features user defined zones, tracks distance travelled, speed and resting time in each zone

**Read**

Connectivity with any video tracking system that can be handled remotely without any human interface and that feature user define zones, tracks distance travelled, speed and resting in each zone