



Date: -15th November, 2017

Corrigendum
For
Tender for
Fully Automated Gel Documentation System for the
Department of Microbiology

NIT Issue Date	: 09 th October, 2017
NIT No.	: Admn/Tender/208/2017-AIIMS.JDH
Pre-Bid Meeting	: 24 th October, 2017 at 03:30 PM
Last Date of Submission	: 27 th November, 2017 at 03:00 PM
Bid opening	: 28 th November, 2017 at 03:15 P.M

The following revised and additional specification will be added:-

1. Page No. 10, Point No. 05:

For

Should have high resolution camera with 4 mega pixel or more/greater resolution, 16-bit CCD Camera, motorized fixed lens 50mm, $f/0.95$

Read

Should have high resolution camera with 4 mega pixel or more/ greater resolution, 16-bit CCD Camera with Peltier Cooling up to -25° C, motorized fixed lens 50mm, $f/0.95$.

2. Page No. 10, Point No. 07:

For

Illumination source should be 306nm trans UV and LED epiwhite. Should have large 10.4 inch touch screen display with an integrated computer with >200GB hard drive. To operate the touchscreen interface, a stylus should be provided along with the system.

Read

Illuminating source with Blue Light Imaging, should be 306nm trans UV and LED epiwhite. Should have large 10.4 inch touch screen display with an integrated computer with >200GB hard drive. To operate the touchscreen interface, a stylus should be provided along with the system.