



Date: -21st December, 2017

Corrigendum
For
Tender for
Horizontal Rectangular High Pressure High Vacuum
Double Door Sterilizer for the Department of CSSD

NIT Issue Date	: 09 th October, 2017
NIT No.	: Admn/Tender/217/2017-AIIMS.JDH
Pre-Bid Meeting	: 23 rd October, 2017 at 03:00 PM
Earlier Last Date of Submission	: 01 st January, 2017 at 03:00 PM
Extended Last Date of Submission	: 11 th January, 2018 at 03:00 PM
Bid opening	: 12 th January, 2018 at 03:15 P.M

The following revised and additional specification will be added:-

1. Page No 10 , Point No 5:

For

System should be European CE certified or USFDA approved and should comply with EN-285 & EN- 554 standards and also confirming to the directives like 93 / 421 EEC (Medical Devices) & 93/23/EC (Pressure equipment directive)

Read

System should be European CE certified or USFDA approved and should comply with EN-285 & EN- 554 standards and also confirming to the directives like 93/42/EEC (Medical Devices & 97/23/EC (Pressure equipment directive)

2. Page No 10, Point No 10, Accessories:

For

- One spare element and two gaskets
- Lead Free Autoclave Steam indicator Tape, one inch, 60 yards long —18 rolls

Read

- One spare element
- Lubricant for Door Seal-15
- Door Gaskets=10
- Bacteria filer for autoclave=10
- Printing paper rolls for autoclave=50
- Lead Free Autoclave Steam indicator tape (1 inch 60 yards) —18 rolls

3. Page No 13, After Point No 10:

Added Point No. 11:

Rate List for accessories should be supplied.