



Date: - 19th October, 2019

Corrigendum
for
Portable Color Doppler Ultrasound Unit for the
Department of Pediatric Surgery

NIT Issue Date	: 07 th August, 2019
NIT No.	: Admn/Tender/23-5/2018-AIIMS.JDH
Pre-Bid Meeting	: 19 th August, 2019 at 05:00 PM
Earlier Last Date of Submission	: 15 th October, 2019 at 03:00 PM
Extended Last Date of Submission	: 30 th October, 2019 at 03:00 PM
Bid opening	: 31 st October, 2019 at 03:15 P.M

The following revised and additional specification will be added:-

1. Page No. 10, Point No 01:

For

The unit must be compact, portable and lightweight, weighing less than 6 kg including battery.

Read as

The unit must be compact, portable and lightweight, weighing less than 6 kg (± 1) including battery.

2. Page No. 10, Point No 13:

For

Flat LCD/ TFT monitor of at least 12 inches with up/down/right/left viewing are >80 degree for wider view angle display in operating rooms.

Read As

Flat LCD/ TFT monitor of at least 15 inches with up/down/right/left viewing are >80 degree for wider view angle display in operating rooms.