



Date: - 29th September, 2020

Corrigendum
for
Portable PTA with Tympanometry for the
Department of ENT

NIT Issue Date	: 12 th June, 2020
NIT No.	: Admn/Tender/27/2020-AIIMS.JDH
Pre-Bid Meeting	: 24 th June 2020 at 03:00 PM
Earlier Last Date of Submission	: 31 st August, 2020 at 03:00 PM
Extended Last Date of Submission	: 15 th September, 2020 at 03:00 PM
Bid opening	: 16 th September, 2020 at 03:15 P.M

The following revised and additional specification will be added:-

1. Chapter I — Sr. No. 1:

For:

Portable PTA with Tympanometry

Read As:

Portable PTA and Portable Tympanometer

2. Annexure- I. Sr. No. 1:

For:

Portable PTA with Tympanometry

Read As:

Portable PTA and Portable Tympanometer

3. Page NO. 10 ,In Technical Specifications, Add following points:

Add

Portable light weight tympanometer with 256 Hz and 1 KHz probe tones, -300 to +300 daPa pressure range, 500, 1000, 2000, 3000, 4000Hz, BB, HP, LP acoustic reflex stimuli, ipsilateral and contralateral reflexes (provide headsets also), touch sensitive display, Internal memory for measurement results. Workflow including tympanometry and reflex, Direct Print of screening results, TOPAE, DPAOE, Data export to a computer.

4. Page NO. 10, Under “Other Requirements”, Add following points:

Add

Other requirements — Add "Both the portable audiometer and portable tympanometer should be from the same OEM and suitable computer software for both should be provided."