



Date: - 04th April, 2020

Corrigendum
for
ICU multipara monitor with central monitoring
system, transport monitor/ module and HL7 for ICU

NIT Issue Date	: 27 th January, 2020
NIT No.	: Admn/Tender/67/2019-AIIMS.JDH
Pre-Bid Meeting	: 06 th February, 2020 at 03:00 PM
Last Date of Submission	: 26 th March, 2020 at 03:00 PM
Last Date of Submission	: 15 th April, 2020 at 03:00 PM
Bid opening	: 16 th April, 2020 at 03:15 P.M

The following revised and additional specification will be added:-

1. Page No. 11, Point No. 16, Addition of Points:

ADD:

Device Integration hub

1. Device Integration hub for integration of devices like syringe pumps. Ventilators and other, at each bed side should be medical grade for ICU charting solution.
2. Device integration hub should be able to connect minimum 4 different devices at each bedside.
3. Ability to connect any device freely to any port of integration hub, without limitation to a specific port. It should automatically detect the equipment without any manual intervention.

2. Page No. 11, Point No. 17:

For

Scope of supply of Hardware:

- b. Laptop with Trolleys: 14 Number
- c. Connection Box: 36 Number

Read as

Scope of Supply of Hardware:

- b. Laptop with Trolleys for ICU: 14 Number & All in one PC for OT & Recovery room: 36 Number
- c. Connection Box: 64 Number

3. Point 11 of corrigendum-I:

For

Post op Beds- 23

Read as:

1. Post op Beds- 12