Tender

For

Advanced Dental Chair

At

All India Institute of Medical Sciences, Jodhpur

NIT No. : Admin/General/42/2013-AIIMS.JDH.

NIT Issue Date : June 07, 2013.

Last Date of Submission : June 17, 2013 at 3:00 PM.



All India Institute of Medical Sciences, Jodhpur

Basni Phase - II, Jodhpur — 342005, Rajasthan Telefax: 0291- 2740329, email: <u>aoadmin@aiimsjodhpur.edu.in</u> <u>www.aiimsjodhpur.edu.in</u> All India Institute of Medical Sciences (AIIMS), Jodhpur, Rajasthan, an apex healthcare institute being established by Parliament of India under aegis of Ministry of Health & Family Welfare, Government of India, invites sealed tenders for supply & installation of the following items at the institute. You are requested to quote your best offer along with the complete details of specifications, terms & conditions.

S.No.	Item Description	Quantity	
1.	Advanced Dental Chair	1	

(Refer Details as per Annexure-'I')

Quotation should be sealed and superscribed with tender number and address to:

"Administrative Officer
All India Institute of Medical Sciences, Jodhpur
Basni, Phase-II
Jodhpur-342005, Rajasthan".

The sealed quotations should reach the Institute, latest by June 17, 2013 at 03:00 PM and it will be opened on same day at 03:45 PM in the Project Cell, Resident Complex, AIIMS, Jodhpur of the Institute in the presence of the bidder(s) or their authorized representative(s), who will present at the scheduled date and time.

Terms & Conditions:

1. Earnest Money Deposit: The bidder shall be required to submit the Earnest Money Deposit (EMD) for an amount of Rs. 10,000/- (Rupees Ten Thousand only) by way of demand drafts only. The demand drafts shall be drawn in favour of "All India Institute of Medical Sciences, Jodhpur". The demand drafts for earnest money deposit must be enclosed in the envelope containing the technical bid.

The EMD of the successful bidder shall be returned after the successful completion of contract / order and for unsuccessful bidder(s) it would be returned after award of the contract. Bid(s) received without demand drafts of EMD will be rejected.

The firms who are registered with National Small Industries Corporation (NSIC) / OR Small Scale Industrial (SSI) are exempted to submit the EMD (copy of registration must be provide along with)

- 2. Rate: Rate should be quoted in Indian Rupees (INR) on DOOR Delivery Basis at AIIMS, Jodhpur, Rajasthan, Inclusive of all the Charges, with break-ups as:
 - Basic Cost.
 - VAT /CST as applicable.
 - Total Cost (F.O.R at AIIMS Jodhpur).

- **3. Validity:** The quoted rates must be valid for a period for 120 days from the date of closing of the tender. The overall offer for the assignment and bidder(s) quoted price shall remain unchanged during the period of validity. If the bidder quoted the validity shorter than the required period, the same will be treated as unresponsive and it may be rejected.
 - In case the tenderer withdraws, modifies or change his offer during the validity period, bid is liable to be rejected and the earnest money deposit shall be forfeited without assigning any reason thereof. The tenderer should also be ready to extend the validity, if required, without changing any terms, conditions etc. of their original tender.
- 4. Warranty / Guarantee: Bid must be quoted with two (02) year comprehensive on-site warranty / guarantee and it will be started from the date of the satisfactory installation / commissioning of goods, against the defect of any manufacturing, workmanship and poor quality of the components. Warranty would be followed by 3 year CMC (Comprehensive maintenance contract with spears) @ 4% of the cost of the Advanced Dental Chair with 5% increment every year. Vendor would submit list of Consumables required for running the machine along with its cost.
- **5. Delivery & Installation**: All the goods ordered shall be delivered & installed within 2 weeks from the date of issuing purchase order. All the aspects of safe delivery, installation and commissioning shall be the exclusive responsibility of the supplier.
 - If the supplier fails to delivered, installation and commissioning of the goods on or before the stipulated date, then a penalty at the rate of 2% per week of the total order value shall be levied subject to maximum of 10% of the total order value.
- **6. Performance Security:** The supplier shall require to submit the performance security in the form of irrevocable Bank Guarantee (BG) / or Fixed Deposit Receipt (FDR) issued by any Nationalised Bank for an amount of which is equal to the 10% of the order value and should be kept valid for a period of 30 day beyond completion of all the contractual obligation, Including CMC period.

7. Payment Term:

- 90% payment of the total order value shall be released after the successful installation/ commissioning of the ordered goods against the submission of the test report.
- Balance 10% of the order value shall be released after the submission of the performance security.
- **8.** Bidder shall submit a copy of the tender document and addenda thereto, if any, with each page of this document should be signed and stamped to confirm the acceptance of the entire terms & conditions as mentioned in the tender enquiry document.

- **9.** Signed & stamped compliance sheet of the technical specification of the goods with technical printed literature must be enclosed with the bid.
- **10.** After due evaluation of the bid(s) Institute will award the contract to the lowest evaluated responsive tenderer
- **11.** Conditional bid will be treated as unresponsive and it may be rejected.
- 12. The Institute reserves the right to accept in part or in full or reject any or more quotation(s) without assigning any reasons or cancel the tendering process and reject all quotations at any time prior to award of contract, without incurring any liability, whatsoever to the affected bidder or bidder(s).

13. Applicable Law:

- The contract shall be governed by the laws and procedures established by Govt.
 of India, within the framework of applicable legislation and enactment made
 from time to time concerning such Commercial dealings / processing.
- Any disputes are subject to exclusive jurisdiction of Competent Court and Forum in Jodhpur, Rajasthan, India only.
- The Arbitration shall be held in accordance with the provisions of the Arbitration and Conciliation Act, 1996 and the venue of arbitration shall be at Jodhpur. The decision of the Arbitrator shall be final and binding on both the partied.
- Force Majeure: Any delay due to Force Majeure will not be attributable to the supplier.

Annexure-I

control on coupling supplied wit 3.4.It should have one high spe with handpiece 3.5.One air motor and micromo contra angle handpieces 3.6.It should have LED light Intensity 800 mW/cm2 and wa output) 3.7.It should have one in-b (frequency 28-36 KHz) with 3 s	cations Qty.
and surgical procedures. 2 Operational Requirements 2.1 Physiological dental chair op 3 Technical Specifications 3.1.Dental unit should have late 3.2.It should have two 3 way sy spare tips) one on unit side and 3.3.It should have one high spec control on coupling supplied wit 3.4.It should have one high spec with handpiece 3.5.One air motor and micromo contra angle handpieces 3.6.It should have LED light Intensity 800 mW/cm2 and wa output) 3.7.It should have one in-b (frequency 28-36 KHz) with 3 s	1
2 Operational Requirements 2.1 Physiological dental chair op 3 Technical Specifications 3.1.Dental unit should have late 3.2.It should have two 3 way sy spare tips) one on unit side and 3.3.It should have one high spec control on coupling supplied wit 3.4.It should have one high spec with handpiece 3.5.One air motor and micromo contra angle handpieces 3.6.It should have LED light Intensity 800 mW/cm2 and wa output) 3.7.It should have one in-b (frequency 28-36 KHz) with 3 s	dental chair required for dental
2.1 Physiological dental chair op 3 Technical Specifications 3.1.Dental unit should have late 3.2.It should have two 3 way sy spare tips) one on unit side and 3.3.It should have one high spec control on coupling supplied wit 3.4.It should have one high spec with handpiece 3.5.One air motor and micromo contra angle handpieces 3.6.It should have LED light Intensity 800 mW/cm2 and wa output) 3.7.It should have one in-b (frequency 28-36 KHz) with 3 s	
2.1 Physiological dental chair op 3 Technical Specifications 3.1.Dental unit should have late 3.2.It should have two 3 way sy spare tips) one on unit side and 3.3.It should have one high spec control on coupling supplied wit 3.4.It should have one high spec with handpiece 3.5.One air motor and micromo contra angle handpieces 3.6.It should have LED light Intensity 800 mW/cm2 and wa output) 3.7.It should have one in-b (frequency 28-36 KHz) with 3 s	
3.1.Dental unit should have late 3.2.It should have two 3 way sy spare tips) one on unit side and 3.3.It should have one high spec control on coupling supplied wit 3.4.It should have one high spec with handpiece 3.5.One air motor and micromo contra angle handpieces 3.6.It should have LED light Intensity 800 mW/cm2 and wa output) 3.7.It should have one in-b (frequency 28-36 KHz) with 3 s	erated by electricity
3.1.Dental unit should have late 3.2.It should have two 3 way sy spare tips) one on unit side and 3.3.It should have one high spec control on coupling supplied wit 3.4.It should have one high spec with handpiece 3.5.One air motor and micromo contra angle handpieces 3.6.It should have LED light Intensity 800 mW/cm2 and wa output) 3.7.It should have one in-b (frequency 28-36 KHz) with 3 s	
3.2.It should have two 3 way sy spare tips) one on unit side and 3.3.It should have one high speed control on coupling supplied with 3.4.It should have one high speed with handpiece 3.5.One air motor and micromode contrated angle handpieces 3.6.It should have LED light Intensity 800 mW/cm2 and was output) 3.7.It should have one in-background (frequency 28-36 KHz) with 3 sections.	st overhead delivery system
spare tips) one on unit side and 3.3.It should have one high speed control on coupling supplied with 3.4.It should have one high speed with handpiece 3.5.One air motor and micromode contrate angle handpieces 3.6.It should have LED light Intensity 800 mW/cm2 and was output) 3.7.It should have one in-background (frequency 28-36 KHz) with 3 seeps to speed the should have one in-background the same should be supplied to should have one in-background the same should be supplied to sho	* *
3.3.It should have one high speed control on coupling supplied with 3.4.It should have one high speed with handpiece 3.5.One air motor and micromode contrate angle handpieces 3.6.It should have LED light Intensity 800 mW/cm2 and was output) 3.7.It should have one in-background (frequency 28-36 KHz) with 3 seeps to supplie the control of the same should be supplied to the control of the same should be supplied to the same should be	
3.4.It should have one high spewith handpiece 3.5.One air motor and micromocontra angle handpieces 3.6.It should have LED light Intensity 800 mW/cm2 and waoutput) 3.7.It should have one in-be (frequency 28-36 KHz) with 3 seconds	ed Air Rotor terminal with water
with handpiece 3.5.One air motor and micromo contra angle handpieces 3.6.It should have LED light Intensity 800 mW/cm2 and wa output) 3.7.It should have one in-b (frequency 28-36 KHz) with 3 s	h handpieces.
3.5.One air motor and micromo contra angle handpieces 3.6.It should have LED light Intensity 800 mW/cm2 and wa output) 3.7.It should have one in-b (frequency 28-36 KHz) with 3 s	ed fiber-optic air-rotor terminal
contra angle handpieces 3.6.It should have LED light Intensity 800 mW/cm2 and wa output) 3.7.It should have one in-b (frequency 28-36 KHz) with 3 s	
3.6.It should have LED light Intensity 800 mW/cm2 and wa output) 3.7.It should have one in-b (frequency 28-36 KHz) with 3 s	tor terminal having straight and
Intensity 800 mW/cm2 and wa output) 3.7.It should have one in-b (frequency 28-36 KHz) with 3 s	cure unit on unit sides (Min
output) 3.7.It should have one in-b (frequency 28-36 KHz) with 3 s	•
(frequency 28-36 KHz) with 3 s	
	ouilt Piezon Ultrasonic Scaler
	caler tips and one set of perio-
curette tips	
pads. 6 spare autoclavable pads	nould be kept on Autoclavable
3.9 Removable auxillary tray (sta	
3.10.It should have latest LED/h	
	low vacuum motorized with
continuous nonstop function suc	ction
	le programmable and return to
Zero position	
3.13It should have LED based X-	•
3.14. It should be provided with	ional foot control base (fixed or
mobile)	onal root control base (fixed of
	ith one doctor's stool and one
assistant's stool with adjusta	ble backrest tilt including an
adjustable ring for foot rest.	
	L HP with silicon filter and dryer
(Medical Grade)	itan manustina Anna
3.18 Should have attached moni	_
4 System Configuration Accesso 4.1 System as specified	ries, spares and consumables

- 4.2 All consumables required for installation and standardization of system to be given free of cost.
- 4.3 Provision for modular furniture with cabinets for storage of equipment.
- 5 Environmental factors
- 5.1 The unit shall be capable of being stored continuously in ambient temperature of 0 -50deg C and relative humidity of 15-90%
- 5.2 The unit shall be capable of operating continuously in ambient temperature of 10-40deg
- 5.3 Complete installation of the system including water input and drainage system has to be installed
- 6 Power Supply
- 6.1 Power input to be 220-240VAC, 50Hz
- 6.2 Five KV Servo Voltage stabilizer of appropriate ratings meeting ISI Specifications.(Input 160-260 V and output 220-240 V and 50 Hz)
- 7 Standards, Safety and Training
- 7.1 Should be FDA/ CE approved product
- 7.2 Manufacturer/Supplier should have ISO certification for quality standards.
- 7.3 Electrical safety conforms to standards for electrical safety IEC-60601 / IS-13450.
- 8 Documentation
- 8.1 User/Technical/Maintenance manuals to be supplied in English.
- 8.2 Certificate of calibration and inspection.
- 8.3 List of important spare parts, handpieces, and accessories with their part number and costing
- 8.4 Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist. The job description of the hospital technician and company service engineer should be clearly spelt out.

Annexure-II

Inviting of sealed quotations for supply and installation of Advanced Dental Chair in the Dentistry of AIIMS, Jodhpur

TECHNICAL BID

Name of Firm/Contractor/Supplier	
Complete Address &	
Telephone No.	
1	
Name of Proprietor/Partner/Managing	
Director/Director.	
Phone & Mobile No.	
Name and address of service centre nearby	
jodhpur.	
Whether the firm is a registered firm	
Yes/No (attached copy of certificate)	
PAN No.	
(enclose the attested copy of PAN Card)	
Service Tax No.	
(enclose the attested copy of Service Tax	
Certificate)	
VAT No.	
(enclose the attested copy of VAT	
Certificate)	
Whether the firm has enclosed the Bank	
Draft/Pay Order/Banker's cheque of	
Rs.10,000/- as Earnest Money Deposit.	
Whether the Firm/Agency has signed each	
and every page of Tender/NIT	
Please provide full list consumable.	
Any other information, if necessary	

Authorized signatory of the bidder with seal.

Annexure-III

Format for Financial Bid

(To be submitted on the letterhead of the company / firm)

S.No	Name of Item	Quantity	Rate	Vat/Taxes	Amount
1.	Advanced Dental Chair	1			
	Grand Total				

- 1. I/We have gone through the terms & conditions as stipulated in the tender enquiry document and confirm to accept and abide the same.
- 2. No other charges would be payable by the Institute.

Authorized signatory of the bidder with seal.