



Date: -14<sup>th</sup> July, 2017

Corrigendum  
For  
Tender for  
Blood and Fluid Warmer for the Department of  
Pediatrics

NIT Issue Date	: 11 <sup>th</sup> May, 2017
NIT No.	: Admn/Tender/108/2017-AIIMS.JDH
Pre-Bid Meeting	: 23 <sup>rd</sup> May, 2017 at 03:30 PM
Earlier Last Date of Submission	: 19 <sup>th</sup> July, 2017 at 03:00 PM
Extended Last Date of Submission	: 01 <sup>st</sup> August, 2017 at 03:00 PM
Bid opening	: 02 <sup>nd</sup> August, 2017 at 03:30 P.M

**The following revised and additional specification will be added:-**

- 1. Page No. 10, Annexure I, S No. 1, Item: Blood and Fluid warmer, Point No. 2**  
**For**  
Should be able to maintain or warm fluid/blood at a flow rate of 2.5 L/min  
**Read**  
Should be able to maintain or warm fluid/blood at a flow rate of **5-100 mL/min**
- 2. Page No. 10, Annexure I, S No. 1, Item: Blood and Fluid warmer, Point No. 3**  
**For**  
Should have a digital temperature display of fluid  
**Read**  
Should have a digital temperature display or **Numerical display** of fluid
- 3. Page No. 10, Annexure I, S No. 1, Item: Blood and Fluid warmer, Point No. 5**  
**For**  
Should have a warm water column or heated sleeve up to the patient end to maintain the temperature up to the point of entry into the vein.  
**Read**  
Should have a warm water column or heated sleeve **or warming module** up to the patient end to maintain the temperature up to the point of entry into the vein.
- 4. Page No. 10, Annexure I, S No. 1, Item: Blood and Fluid warmer, Point No. 8**  
**For**  
Should be useful for both in adult and Pediatric patients  
**Read**  
Should be useful for both in adult and Pediatric patients **but priming volume should be in a range of 3-5 mL**