



Date: - 05th September, 2019

Corrigendum
for
Tender for
Skull Clamp with brain retractor for the Department of
Neurosurgery

NIT Issue Date	: 17 th June, 2019
NIT No.	: Admn/Tender/144-2/2018-AIIMS.JDH
Pre-Bid Meeting	: 27 th June, 2019 at 04:15 PM
Last Date of Submission	: 29 th August, 2019 at 03:00 PM
Last Date of Submission	: 16 th September, 2019 at 03:00 PM
Bid opening	: 17 th September, 2019 at 03:15 P.M

The following revised and additional specification will be added:-

- 1. Page No. 10, Heading A, starting from “Each Unit/set should have” , Point 1:**
For
Support/Mount device for base unit to any OT table (Original) - 1 No.
Read
Support/Mount device for base unit to any OT table (Original) - 2 No.
- 2. Page No. 10, Heading A, starting from “Each Unit/set should have” , Point 2:**
For
Ultra adjustable Base Unit- 1 No
Read
Ultra adjustable Base Unit- 02 No
- 3. Page No. 10, Heading A, starting from “Each Unit/set should have” , Point 3:**
For
Ball Socket Swivel adaptor- 1 No
Read
Ball Socket Swivel adaptor- 2 No
- 4. Page No. 10, Heading A, starting from “Each Unit/set should have” , Point 4:**
For
Swivel Adaptor – 1 Nos
Read
Deleted
- 5. Page No. 10, Heading A, starting from “Each Unit/set should have” , Point 5:**
For
Skull Clamp- 1 Nos
Read

Skull Clamp- 2 Nos

6. **Page No. 10, Heading A, starting from “Each Unit/set should have” , Point 8:**
For
Adult Horseshoe Head Rest with Extension bar for traction - 1 No.
Read
Adult Horseshoe Head Rest with Extension bar for traction - 2 Nos
7. **Page No. 10, Heading A, starting from “Each Unit/set should have” , Point 9:**
For
Pediatric Horseshoe Head Rest- 1 No
Read
Pediatric Horseshoe Head Rest- 2 Nos
8. **Page No. 10, Heading A, starting from “Each Unit/set should have” , Point 11:**
For
Crossbar Adaptor (Original) for sitting position - 1 No
Read
Crossbar Adaptor (Original) for sitting position - 02 No
9. **Page No. 10, Heading A, starting from “Each Unit/set should have” , Point 12:**
For
Universal Side Rail (Original) - 2 Nos.
Read
Universal Side Rail (Original) - 4 Nos.
10. **Page No. 11, Heading A, starting from “Each Unit/set should have” , Point 15:**
For
Armrest Park-bench = 01
Read
Deleted
11. **Page No. 11, Heading A, starting from “Each Unit/set should have” , Point 16:**
For
Side Rail OR Table Adaptor = 01
Read
Doro Armrest parkbench- 01
12. **Page No. 11, Heading B, starting from “The system configuration for single retractor system.....” :**
For
The system configuration for single retractor system consists of the following:
Read
The system configuration for two retractor system consists of the following:
13. **Page No. 11, Heading B, starting from “The system configuration for single retractor system.....”, Point 1 :**
For
Quick clamp – 02
Read
Quick clamp – 04
14. **Page No. 11, Heading B, starting from “The system configuration for single retractor system.....”, Point 2 :**
For
Coupling clamp – 04
Read
Coupling clamp – 08

- 15. Page No. 11, Heading B, starting from “The system configuration for single retractor system.....”, Point 3 :**
For
Rod -02
Read
Rod -04
- 16. Page No. 11, Heading B, starting from “The system configuration for single retractor system.....”, Point 4 :**
For
Curve rods -02
Read
Curve rods -04
- 17. Page No. 11, Heading B, starting from “The system configuration for single retractor system.....”, Point 5 :**
For
Flexible rotatory arms, 280 mm – 03
Read
Flexible rotatory arms, 280 mm – 06
- 18. Page No. 11, Heading B, starting from “The system configuration for single retractor system.....”, Point 6 :**
For
Retractor blade set 4-inches, black coated for anti-reflective and non-stick with width 6 to 25 mm– 05 blades.
Read
Retractor blade set 4-inches, black coated for anti-reflective and non-stick with width 6 to 25 mm– 20 blades.
- 19. Page No. 11, Heading B, starting from “The system configuration for single retractor system.....”, Point 7 :**
For
Serrated micro-retractor blade with 2 mm × 10mm, length from 4.5’ to 6.5’- 03 nos.
Read
Deleted
- 20. Page No. 11, Heading B, starting from “The system configuration for single retractor system.....”, Point 8 :**
For
Hooked micro-retractor blade with 4mm × 10mm, length from 4.5’ to 6.5’- 03 nos.
Read
Deleted
- 21. Page No. 11, Heading B, starting from “The system configuration for single retractor system.....”, Point 9 :**
For
Separate sterilization trays for retractor and flexible arm- 02 ((1+1)
Read
Separate sterilization trays for retractor and flexible arm- 04 (2+2)
- 22. Page No. 11, Heading B, starting from “The system configuration for single retractor system.....”, Point 10 :**
For
Halo Pattie tray- 01
Read
Halo Pattie tray- 02