



Date: - 21st June, 2018

Corrigendum
For
Tender for
RO Plant (1000 lit/hr) for the Department of
Nephrology

NIT Issue Date	: 09th April, 2018
NIT No.	: Admn/Tender/17/2018-AIIMS.JDH
Pre-Bid Meeting	: 20 th April, 2018 at 03:45 PM
Earlier Last Date of Submission	: 25 th June, 2018 at 03:00 PM
Extended Last Date of Submission	: 04 th July, 2018 at 03:00 PM
Bid opening	: 05 th July, 2018 at 03:15 P.M

The following revised and additional specification will be added:-

1. Page No. 11, 3rd Sub point of Pre Treatment:

For

Raw water tank having food grade quality at least 1000 litres capacity to store raw water.

Read

Raw water tank having food grade quality at least 750 litres or above capacity to store raw water.