



Date: - 14th August, 2019

Corrigendum

For

Fractional Carbon Dioxide (Co2) Laser for the Department of Dermatology, Venereology and Leprology

NIT Issue Date	: 19 th June, 2019
NIT No.	: Admn/Tender/17/2019-AIIMS.JDH
Pre-Bid Meeting	: 03 rd July, 2019 at 03:45 PM
Earlier Date of Submission	: 13 th August, 2019 at 03:00 PM
Extended Date of Submission	: 29 th August, 2019 at 03:00 PM
Bid opening	: 30 th August, 2019 at 03:15 P.M

The following revised and additional specification will be added:-

1. Page No. 02, In technical specification, Point No. 10:

For

The device should have a Sealed LASER tube, DC excited

Read

The device should have a sealed Laser tube, DC/RF excited

2. Page No. 02, In technical specification, Point No. 13:

For

The device should have focused Incisional hand pieces:- 125mm, 200 mm, 260 mm

Read

The device should have focused Incisional Hand pieces of 125mm, 200 mm, 260 mm, atleast

3. Page No. 02, In technical specification, Point No. 13:

For

The device should have option for Deep and Superficial resurfacing in single scan

Read

The device should preferably have option for Deep and superficial resurfacing in single scan