



Date: - 23<sup>rd</sup> January 2021

## Corrigendum-3

for

## Flow Cytometer for the Department of Microbiology

NIT Issue Date	: 30 <sup>th</sup> September, 2020
NIT No.	: Admn/Tender/36-3/2020-AIIMS.JDH
Pre-Bid Meeting	: 12 <sup>th</sup> October 2020 at 04:00 PM
Earlier Last Date of Submission	: 28 <sup>th</sup> January, 2020 at 03:00 PM
Extended Last Date of Submission	: 09 <sup>th</sup> February, 2021 at 03:00 PM
Bid opening	: 10 <sup>th</sup> February, 2021 at 03:15 P.M

**The following revised and additional specification will be added: -**

**1. In Corrigendum dated 21<sup>st</sup> December 2020, Page No. 02, Point No. 5:**

**For**

The system should have high quality stainless steel/quartz flow cell & system should have fluorescence detectors of PMT bases.

**Read as**

The system should have high quality stainless steel/quartz flow cell & system should have fluorescence detectors of PMT bases/APDs.