



Date: -21<sup>st</sup> June, 2017

Corrigendum  
For  
Tender for  
Ultrasonic Aspirator for use with Endoscopic  
Endonasal Surgeries for the Department of ENT

NIT Issue Date	: 03 <sup>rd</sup> April, 2017
NIT No.	: Admn/Tender/45/2017-AIIMS.JDH
Pre-Bid Meeting	: 18 <sup>th</sup> April, 2017 at 03:45 PM
Earlier Last Date of Submission	: 21 <sup>st</sup> June, 2017 at 03:00 PM
Extended Last Date of Submission	: 10 <sup>th</sup> July, 2017 at 03:00 PM
Bid opening	: 11 <sup>th</sup> July, 2017 at 03:30 P.M

**The following revised and additional specification will be added:-**

- 1. Page No. 11, In technical specification, Para No. 03, In Ultrasonic Aspiration, Sub Point No. 1:**  
**For**  
Piezoelectric transducer, power to size ratio, Longitudinal & Torsional Vibration Technology for bone dissection in skull base, spine with Nonrotational reciprocating motion and should be compatible with Universal 25 KHz hand piece.  
**Read**  
Piezoelectric transducer, power to size ratio, Longitudinal & Torsional Vibration Technology for bone dissection in skull base, with Nonrotational reciprocating motion and should be compatible with Universal 22 - 25KHz hand piece.
- 2. Page No. 11, In technical specification, Para No. 04, In Console, Sub Point No. 04:**  
**For**  
Operator control: Full Touch panel  
**Read**  
Operator control: Full Touch/ Touch pad panel
- 3. Page No. 11, In technical specification, Para No. 04, In Console, Sub Point No. 11:**  
**For**  
Weight of console should be less than 24 Kg  
**Read**  
Console should be provided with the trolley or console inbuilt with trolley

- 4. Page No. 11, In technical specification, Para No. 04, In Console, Sub Point No. 12:**  
**For**  
Single Pedal Foot Switch Cord Length should be 12 feet  
**Read**  
Multifunctional Wireless footswitch
- 5. Page No. 11, In technical specification, Para No. 05, In Alarms, In Heading:**  
**For**  
Should have following alarms  
**Read**  
Should have following alarms with Error code to troubleshoot the error:
- 6. Page No. 11 and 12, In technical specification, Para No. 06, In Hand piece, Para No. 01:**  
**For**  
Universal handpiece for soft tissue & bone dissection having various compatible tips from soft tissue to bone to transphenoidal length.  
**Read**  
Universal short straight handpiece for soft tissue & bone dissection having various compatible tips stright and curved from soft tissue to bone to transphenoidal length. If vendor/ manufacturer don't have curved tip should be provide Straight and curved Handpiece for straight tips.
- 7. Page No. 12, In technical specification, Para No. 06, In Hand piece, Para No. 02:**  
**For**  
Lightweight, 120 grams, ergonomic design provides superb tactile feedback and requires no external cooling  
**Read**  
Lightweight, 70- 120 grams, ergonomic design provides superb tactile feedback and requires no external cooling for long working hors
- 8. Page No. 12, In technical specification, Para No. 06, In Hand piece, Para No. 03:**  
**For**  
25 kHz Universal Hand piece  
**Read**  
22-25 kHz Universal Hand piece
- 9. Page No. 12, In technical specification, Para No. 06, In Hand piece, Para No. 04:**  
**For**  
14 types of specialty tips should be attachable  
**Read**  
Minimum 11 types of tips should be attachable
- 10. Page No. 12, In technical specification, Para No. 06, In Hand piece, Para No. 05:**  
**For**  
Frangible tip covers for improved visualization and access  
**Read**  
Frangible / Silicon tip covers for improved visualization and access
- 11. Page No. 12, In technical specification, Para No. 07, In Tips, Para No. 03:**  
**For**  
At least 14 compatible different tips ...ranging from soft tissue, to bone, to transsphenoidal length  
**Read**  
At least 11 compatible different tips (straight and curved) ...ranging from soft tissue, to bone, to transsphenoidal length. All tips should be provided with system as per our choice.

**12. Page No. 12, In technical specification, Para No. 07, In Tips, Para No. 10:**

**For**

V shape Tip end

**Read**

V shape Tip / Cylindrical shape tip end

**13. Page No. 12, In technical specification, Para No. 07, In Tips, Para No. 13:**

**For**

Two sizes, tip diameter should be available available

**Read**

3 sizes, tip diameter should be available available

**14. Page No. 12, In technical specification, Para No. 07, In Tips, Para No. 14:**

**For**

Ring Curette Tip

**Read**

Ring Curette / Nothed Tip

**15. Page No. 12, In technical specification, Para No. 08, In The System should have below accessories, Para No. 5:**

**For**

Disposable Tubing Set & Extender Filter Tubing

**Read**

At least 25 Nos Aspiration Tubing set should be provided with the system.

(With Extender Filter Tubing)

**16. Page No. 12, In technical specification, Para No. 08, In The System should have below accessories, Para No. 6:**

**For**

Disposable Suction Canister Liner

**Read**

Disposable Suction Canister Liner (desirable)