



Date: - 26<sup>th</sup> February 2019

## Corrigendum

For

### ICU Multipara Monitor with central monitoring system transport Monitor/ module HL7 for ICU

<b>NIT Issue Date</b>	<b>: 27<sup>th</sup> January 2019</b>
<b>NIT No.</b>	<b>: Admn/Tender/67/2019-AIIMS.JDH</b>
<b>Pre-Bid Meeting</b>	<b>: 06<sup>th</sup> February, 2020 at 03:00 PM</b>
<b>Last Date of Submission</b>	<b>: 12<sup>th</sup> March, 2020 at 03:00 PM</b>
<b>Bid opening</b>	<b>: 13<sup>th</sup> March, 2020 at 03:15 PM</b>

**The following revised and additional specification will be added:-**

**1. Page No. 10, In technical specification, Point No. 04:**

**For**

Monitor should be upgradable with Cardiac output, PiCCO, BIS, NMT, AGM, EEGX4, ScVO<sub>2</sub>, & RM through External plug and play module. The monitor should be able to display at least 5 of these parameters simultaneously. Price should be quoted of each module separately.

**Read**

Monitor should be upgradable with Cardiac output, PiCCO, BIS, NMT, AGM, EEGX4, & ScVO<sub>2</sub> through External plug and play module. The monitor should be able to display at least 5 of these parameters simultaneously. Price should be quoted of each module separately.

**2. Page No. 10, In technical specification, Point No. 05:**

**For**

Monitor should have capability of simultaneously monitoring of Min 12 channel of waveform.

**Read**

Monitor should have capability of simultaneously monitoring of Min 8 channel of waveform.

**3. Page No. 10, In technical specification, Point No. 08:**

**For**

Monitor should have non-volatile graphic and tabular trending of all monitored parameters as standard for minimum 120 hrs.

**Read**

Monitor should have non-volatile graphic and tabular trending of all monitored parameters as standard for minimum 72 hrs.

**4. Page No. 10, In technical specification, Point No. 09:**

**For**

Monitor should have facility to record of 48 ECG full disclosure

**Read**

Monitor should have facility to record of 48 hours or more full disclosure through central station

**5. Page No. 10, In technical specification, Point No. 11:**

**For**

Monitor should have battery backup for 120 min

**Read**

Monitor should have inbuilt /UPS battery backup for 120 min. The cost of UPS for each monitor to be quoted separately.

**6. Page No. 10, In technical specification, Point No. 12:**

**For**

System should have facility of Remote viewing of data of each monitor on Mobile / Laptop/ Desktop without using any special server.

**Read**

System should have facility of Remote viewing of data of each monitor on Laptop/ Desktop.

**7. Page No. 10, In technical specification, Point No. 14:**

**For**

Six Central Stations:

- a. 22 inch touch screen display
- b. Minimum 16 beds view to be provided in each station
- c. UPS with 15min backup
- d. Dual display option (based on number of monitors to be connected)
- e. Networking cost to be included

**Read**

Three Central Stations:

- a. 22 inch touch screen display
- b. Minimum 16 beds view to be provided in each station
- c. UPS with 15min backup
- d. Dual display option (based on number of monitors to be connected)
- e. Networking cost to be included
- f. Rate of Central station for 16 beds to be quoted separately.

**8. Page No. 11, In technical specification, Point No. 17:**

**For**

Scope of Supply of Hardware:

- a. Server: 1 Number
- b. Laptop with Trolleys: 14 Number
- c. Connection BOX: 44 Numbers
- d. WIFI AP: 6 Numbers
- e. Switch: 4 Numbers
- f. Unit Price of Each item software and Hardware to be Quoted Separately.

**Read**

Scope of Supply of Hardware:

- a. Server: 2 Number
- b. Laptop with Trolleys: 14 Number
- c. Connection BOX: 36 Numbers
- d. WIFI AP: 6 Numbers
- e. Switch: 4 Numbers
- f. Unit Price of Each item software and Hardware to be Quoted Separately.
- g. 24 U Rack Server to be provided
- h. Suitable UPS with minimum 10 minutes backup for all servers

**9. Page No. 11, In technical specification, Point No. 18:**

**For**

All components should be quote from the same manufacturer.

**Read**

Charting and Monitoring solution should be quote from the same manufacturer.

**10. Page No. 11, In technical specification, Add Point No. 19 after Point No. 18:**

**Added**

It should do scoring as per international standard (SOFA, APACHE) and also help to configure new own scoring for study purpose in system.

**11. Page No. 11, In technical specification, Add Point No. 20 after Point No. 19:**

**Added**

Total No of

- a. OTs- 30
- b. Post op Beds- 23