



Dated: - 25<sup>th</sup> October 2016

## Addendum

for

### Equipments required for Department of Radiology

NIT Issue Date	:	05 <sup>th</sup> August 2016.
NIT No.	:	Admn/Tender/Radiology/2016-AIIMS.JDH
Pre Bid Meeting held on	:	16 <sup>th</sup> August, 2016 at 03:15 PM
Last Date of Submission	:	01 <sup>st</sup> September, 2016 at 03:00 PM
Extended Last Date of Submission	:	07 <sup>th</sup> November, 2016 at 03:00 PM
Revised Last Date of Submission	:	15 <sup>th</sup> November, 2016 at 03:00 PM

#### 1. The following revised and additional specification will be added:-

1. **Corrigendum Page No. 02, S.No. 2, 1. GENERATOR and Electrodes, 3. MISCELLANEOUS, Point no. b:**

**For**

10 electrodes should be supplied along with the main equipment-type of electrodes (bone/solid viscera use) will be made after finalization of the L1 bidder. The vendor should also quote the price of each such electrode which would be frozen for the next 5 years from the date of installation of the RF generator.

**Read**

20 electrodes should be supplied along with the main equipment-type of electrodes (bone/solid viscera use) will be made after finalization of the L1 bidder. The vendor should also quote the price of each such electrode which would be frozen for the next 5 years from the date of installation of the RF generator.

2. **Page No. 10, S.No. 2, Radiofrequency Ablation System, 1. GENERATOR and Electrodes :**

**Added Point No. m:**

The equipment should be equipped with saline infusion technology. The pump for the same must be provided along with the standard equipment.

3. **Page No. 10, S.No. 2, Radiofrequency Ablation System, 3. MISCELLANEOUS, Added Point No. n:**

The equipment must be usable in pediatric patients with appropriate size pediatric earthing pads available as standard.

**Administrative Officer  
AIIMS, Jodhpur**