अखिल भारतीय आयुर्विज्ञान संस्थान, जोधपुर
ALL INDIA INSTITUTE OF MEDICAL SCIENCES, JODHPUR

**Date: - 03rd December 2016**

# Corrigendum

# for

# Phaco Machine for the Department of Ophthalmology

| | | |
|---|---|---|
| NIT Issue Date | : | 06th September 2016. |
| NIT No. | : | Admn/Tender/47/2016-AIIMS.JDH |
| Pre Bid Meeting held on | : | 20th September, 2016 at 04:00 PM |
| Last Date of Submission | : | 07th October, 2016 at 03:00 PM |
| Extended Last Date of Submission | : | 12th December, 2016 at 03:00 PM |
| Revised Last Date of Submission | : | 22nd December, 2016 at 03:00 PM |

## 1. The following revised and additional specification will be added:-

1. **Page No. 10, Technical Specification, B. Control, Point No. 3 :**
   **For**
   Wireless remote control
   **Read**
   Wired/ Wireless remote control.

2. **Page No. 10, Technical Specification, B. Control, Point No. 4 :**
   **For**
   Individual programming with ParaProg for up to 50 surgeons
   **Read**
   Individual programming with ParaProg for up to 10 and above surgeons

3. **Page No. 10, Technical Specification, C. In Head Line:**
   **For**
   I/A Function
   **Read**
   I/A Function (Present in almost all Phaco Machine)

4. **Page No. 10, Technical Specification, D. Phaco Function, Point No. 3:**
   **For**
   Hexadisq handpiece with 6 piezo plates
   **Read**
   Hexadisq handpiece with 4 and above piezo plates required

5. **Page No. 10, Technical Specification, D. Phaco Function, Point No. 8:**
   **For**
   Override function
   **Read**
   Override function required (To manually change the preset parameters)

6. **Page No. 10, Technical Specification, F. Bipolar Functions:**
   **For**

| F | Bipolar Functions |
|---|---|
| 1 | Endo-diathermy |
| 2 | Macro-diathermy |
| 3 | Conjunctival coaptation |
| 4 | Instant diathermy function |
| 5 | RF capsulotomy |
| 6 | STT glaucoma sclerothalamotomy |
| 7 | IDK glaucoma intrastromal diathermal keratostomy |

   **Read**

| F | Bipolar Functions |
|---|---|
| 1 | Diathermy with 2 probes required |
| 2 | Instant diathermy function required |
| 3 | **Preferably** RF capsulotomy with or without |
| 4 | Preferably STT glaucoma sclerothalamotomy with or without |
| 5 | Preferably IDK glaucoma intrastromal diathermal keratostomy with or withour. |

**Administrative Officer**
**AIIMS, Jodhpur**