

Invitation of quotation
for
Supply of Equipments for Human Milk Bank
At
All India Institute of Medical Sciences, Jodhpur

Inquiry No.: : Admin/Gen/72-10(ii)/2022-AIIMS.JDH

Inquiry Issue Date : 05th August, 2022

Last Date of Submission : 09th August, 2022 at 03:00 PM.



All India Institute of Medical Sciences, Jodhpur

Basni Phase - II, Jodhpur – 342005, Rajasthan

Telefax: 0291- 2740741, email: procurement@aiimsjodhpur.edu.in

www.aiimsjodhpur.edu.in

Invitation of quotation for Supply of Equipments for Human Milk Bank at AIIMS Jodhpur

Sealed Quotations are hereby invited by the undersigned on behalf of the Director, AIIMS Jodhpur for Supply of Equipments for Human Milk Bank for the Institute as per terms & conditions mentioned below. The filled quotations along with all the required document must reach in the office of the undersigned on or before 09.08.2022 03:00 PM. The Envelope containing the quotation would please be sealed and super scribed as under:-

“QUOTATION FOR SUPPLY OF EQUIPMENTS FOR HUMAN MILK BANK AGAINST INQUIRY NO. ADMN/GEN/72-10(ii)/2022-AIIMS.JDH” DUE ON 09.08.2022 03:00 PM

1. Terms & Conditions:

- A) The quotations received after this deadline & unsealed shall not be entertained under any circumstances whatsoever. In case of postal delay this Institute will not be responsible. **The offer Submitted Fax/Email shall not be considered and no correspondence will be entertained in this matter.**
- B) Quotations must be in the enclosed prescribed Performa on the letter head of the firm duly signed by the Proprietor/ Partner/ Director or their authorized representative, In case of signing of quotation by the authorized representative letter of authorization must be attached with the quotation. Quotation must be dropped in “**Quotation Box**” located in Administration Block of AIIMS, Jodhpur.
- C) Rates must be quoted in **Indian rupees** and as per the format specified taxes extra if any must be written separately.
- D) Rates must be quoted FOR basis (including Freight charges, Insurance, installation etc.)
- E) No overwriting or cutting is permitted in the rate. If found, the quotation shall be summarily rejected.
- F) The rates quoted must be valid for 60 days minimum from the date of opening of the quotation and silence of any tendered on this issue shall be treated as agreed with this condition.
- G) Becoming L1 will not be the criteria for awarding of purchase order unless the rates are reasonable & justified.
- H) RTGS/NEFT details need to be furnished by the supplier with the quotation on the letter head of supplier/firm/agency.
- I) The firm/agency may satisfy the following conditions and attach self-attested copy of the same with the quotation:
- Firm shall be registered with the Government of Rajasthan / Central Government.
 - The firm shall have valid GST No.
 - **The firm should not be black listed by any Govt. Agency/Dept.**
- J) Quotations qualified by such vague and indefinite expressions such as “subject to prior

INQUIRY NO. Admin/Gen/72-10(ii)/2022-AIIMS.JDH

confirmation”, “subject to immediate acceptance” etc. will be treated as vague offers and rejected accordingly. Any conditional quotation shall be rejected summarily.

- K) **Delivery Period** – within 30 days from Purchase order.
- L) **Liquidated Damage:** - If the supplier fails to deliver the material on or before the stipulated date, then a penalty at the rate of 0.5 % per week of the total order value shall be levied subject to maximum of 10% of the total order value.
- M) **Payment Terms:** Payment will be only after satisfactorily delivery / commissioning of material and after inspection by the AIIMS Jodhpur.
- N) **Disputes:** -In the event of any dispute or disagreement arising between the contractors and any other department of AIIMS Jodhpur with regards to the interpretation of “Terms & Conditions” of this inquiry, the same shall be referred to the Director, AIIMS Jodhpur whose decision will be final and binding upon the contractor.
- O) AIIMS, Jodhpur reserves the right to increase or decrease quantity and / or amount of work. Decision of Quantity of material in the AIIMS, Jodhpur will be final in this regard.
- P) AIIMS, Jodhpur reserves the right to reject any quotation or part or the whole of inviting quotation process without assigning any reason. Decision of the AIIMS, Jodhpur will be final in this regard.

2. Special Terms & Conditions:

- A) **Bidder must quote the product as per specification provided in Annexure 1.**
- B) **Catalog must be attached with quotation for technical evaluation.**
- C) **The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.**

Deputy Director (Admin)

Encl.: Annexure 1 (Specification)
Annexure 2 (Format of price bid)

Annexure 1

S. No.	Particular	Specifications	Qty.
1.	Microscope with oil immersion lens	<p><u>Specification of Compound Binocular Microscope: 1 NOS.</u></p> <ol style="list-style-type: none"> 1. Optical system Universal Infinity Corrected Optical system. 2 2. Illuminator Built-in transmitted LED illumination with Co-axial Coarse & fine focusing mechanism with low drive handle. 3. Focusing Stage height movement (coarse movement stroke: 15mm or better), Torque adjustment for coarse focusing knob, Coarse & fine focus knob should be on both sides of the microscope body for easier working. Microscope should have an upper limit stopper for high magnification objectives to secure objective hitting the slide. This is essential requirement. 4. Nosepiece Inward facing quadruple nosepiece. 5. Stage Wire movement mechanical fixed stage (W x D): 174mm x 89mm. Traveling range (X *Y): 76mm X 30mm. 6. Observation Tube Siedentop type binocular observation tube inclined at 30° with field no. 20mm or, more, Interpapillary distance 48-75 mm. The observation tube should have and eyepoint height adjustment system to accommodate user's height for comfort 7. Eyepiece Paired 10X eyepiece with dioptr adjustment with field no. 20mm or more. The eyepiece can be fixed in place to prevent damage/loss when moved 8. Condenser Abbe condenser (N.A 1.25 or higher) with built-in-aperture iris Diaphragm for bright field and dark field. 9. Objectives Completely field Plan achromat objectives 4 X N.A. 0.10 & WD 27.8mm or better 10X N.A.0.25 & WD 8.0mm or better 40 X N. A 0.65 & WD 0.6mm or better 100 X Oil N.A 1.25 & WD 0.13mm or better 10. Important to Note Microscope should essentially be (USFDA/CE Complied. A Suitable certificate in this regard should be supplied. On site demonstration should be done Polarizer supplied with microscope is desirable. 	01 Nos.
2.	Hot air circulating oven	<ol style="list-style-type: none"> 1. Precise Hot Air Drying & Sterilization. SS 304 Grade Stainless Steel. 2. Better Mineral/Glass wool/Blanket Insulation. 3. Silent Hot Air Blower, unique Design of Air Circulation provide throughout uniform air movement & improved Temperature distribution. 4. Unique design of Ventilation keeps the surface of the instrument from being burnt even when the instrument i.e. Oven Temperature maintained at above 200° C. 5. Polish 304 grade S. Steel interior, corrosion resistant long-life operation having SS adjustable shelves to make more space quality coil shaped heaters duly embedded/ wound on the sides & bottom for better accuracy. 	02 Nos.

INQUIRY NO. Admin/Gen/72-10(ii)/2022-AIIMS.JDH

		<ol style="list-style-type: none">6. Cap. 224lit Inner chamber.7. Interior Size: Stainless Steel: 604x604x604mm Cap. 225lit.8. Heater Wattage “W” 2.5KW.9. Shelves made of stainless Steel 2 pcs.10. Control Panel with Mains Load Indicator Temp. Controller Fan11. Temperature Range: Ambient to 200°C controlled from 50°C to 200°C.12. Temperature Stability: $\pm 1^\circ\text{C}$ or Better, In Microprocessor Based PID Temperature Controller $\pm 0.5^\circ\text{C}$ or better.13. Power Consumption 220/230V AC Supply Single Phase 50Hz.	
3.	Reusable Breast Milk Storage Bottles	<ol style="list-style-type: none">1. Bottles made of polypropylene, without BPA and DEHP2. screw cap3. safe for Dishwasher with Locking Ring, Cap4. should be proven food-grade Material.5. Capacity: 250 ml	100 Nos.
4.	Cap sealing machine	-----	01 Nos.
5.	Plastic Bottle Sterilizer	-----	02 Nos.
6.	Bottle sealer with Foil for sealing bottles	-----	01 Nos.
7.	Printer and water proof labels	-----	01 Nos.

Note: -

- ❖ **Catalog/ Brochure must be attached for Technical Evaluation.**
- ❖ The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.

[On the letterhead of firm]

ANNEXURE "2"
PRICE BIDFORM

To,

Deputy Director (Admin),
AIIMS, Jodhpur.

Dear Sir,

1. I/We Submitted the quotation for Enquiry No. "QUOTATION FOR SUPPLY OF EQUIPMENTS FOR HUMAN MILK BANK AT AIIMS AGAINST THE INQUIRY NO. Admn/Gen/72-10(ii)/2022-AIIMS.JDH" DUE ON 09.08.2022 03:00 PM for Supply of Equipments for Human Milk Bank at AIIMS Jodhpur".
2. I/We thoroughly examined, understood and accepted terms & conditions given in the enquiry document, failing which my quotation will be rejected out rightly.
3. I/We hereby offer to supply at the following rates.

S. No	Particular	Qty.	Quoted Make	Price/Unit Exclusive of GST (INR)	GST/Other Taxes	Price/ Unit Inclusive of GST (INR)	Total Cost Inclusive of GST (INR)	MRP
1.	Microscope with oil immersion lens Specification:- As per annexure – 1	01 Nos.						
2.	Hot air circulating oven Specification:- As per annexure – 1	02 Nos.						
3.	Reusable Breast Milk Storage Bottles Specification:- As per annexure – 1	100 Nos.						
4.	Cap sealing machine Specification:- As per annexure – 1	01 Nos.						
5.	Plastic Bottle Sterilizer Specification:- As per annexure – 1	02 Nos.						
6.	Bottle sealer with Foil for sealing bottles Specification:- As per annexure – 1	01 Nos.						
7.	Printer and water proof labels Specification:- As per annexure – 1	01 Nos.						

Note:-

1. **The Bidder must quote only single Make & Model.**
2. The bidder must quoted their quotation only in above said format on the letter of firm otherwise quotation will be **REJECTED**.
3. Catalog must be attached with quotation for technical evaluation.
4. The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.

INQUIRY NO. Admin/Gen/72-10(ii)/2022-AIIMS.JDH

Date _____

Place _____

(Name) _____

Name of Firm/Company/Agency _____

GSTIN No.: _____

Bank Name:- _____

Bank Account No.: _____

IFSC Code:- _____

Branch Name: _____

Phone No. _____

Email: _____

(Signature of Authorized Person) _____

Seal: _____